

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
 Data File : VD070274.D
 Acq On : 31 Aug 2021 12:35
 Operator : VA/SY
 Sample : VD0831SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0831SBL01

Quant Time: Sep 01 03:57:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	494544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	1018937	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	1019350	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	451947	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	375731	64.596	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.200%
35) Dibromofluoromethane	7.914	113	341096	51.815	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.620%
50) Toluene-d8	10.332	98	1292372	50.475	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.940%
62) 4-Bromofluorobenzene	12.626	95	465851	55.239	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	110.480%
Target Compounds						
16) Acetone	4.156	43	15504	14.469	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
Data File : VD070274.D
Acq On : 31 Aug 2021 12:35
Operator : VA/SY
Sample : VD0831SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0831SBL01

Quant Time: Sep 01 03:57:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

