

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
 Data File : VD070278.D
 Acq On : 31 Aug 2021 14:30
 Operator : VA/SY
 Sample : M3605-02
 Misc : 4.68G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PX-2

Quant Time: Sep 01 03:58:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	460448	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	935801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	933519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	401172	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	303968	56.128	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.260%
35) Dibromofluoromethane	7.914	113	304434	50.354	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.700%
50) Toluene-d8	10.338	98	1201340	51.088	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.180%
62) 4-Bromofluorobenzene	12.626	95	406265	52.453	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
Data File : VD070278.D
Acq On : 31 Aug 2021 14:30
Operator : VA/SY
Sample : M3605-02
Misc : 4.68G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PX-2

Quant Time: Sep 01 03:58:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

