

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
 Data File : VD070279.D
 Acq On : 31 Aug 2021 14:59
 Operator : VA/SY
 Sample : M3605-03
 Misc : 5.34G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PX-3

Quant Time: Sep 01 03:58:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
 Quant Title : SW846 8260
 QLast Update : Sat Aug 14 02:33:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	458653	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	943396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	944291	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	408120	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	313475	58.110	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.220%
35) Dibromofluoromethane	7.914	113	308317	50.586	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.180%
50) Toluene-d8	10.338	98	1204897	50.827	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.660%
62) 4-Bromofluorobenzene	12.626	95	409778	52.481	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
Data File : VD070279.D
Acq On : 31 Aug 2021 14:59
Operator : VA/SY
Sample : M3605-03
Misc : 5.34G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PX-3

Quant Time: Sep 01 03:58:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Quant Title : SW846 8260
QLast Update : Sat Aug 14 02:33:51 2021
Response via : Initial Calibration

