



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD083121\
Data File : VD070272.D
Acq On : 31 Aug 2021 08:54
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D081321S.M
Title : SW846 8260
Last Update : Sat Aug 14 02:33:51 2021



AutoFind: Scans 1819, 1820, 1821; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	70587	PASS
75	95	30	60	50.1	197803	PASS
95	95	100	100	100.0	394795	PASS
96	95	5	9	6.7	26512	PASS
173	174	0.00	2	1.3	3725	PASS
174	95	50	100	71.7	283093	PASS
175	174	5	9	7.6	21440	PASS
176	174	95	101	97.4	275627	PASS
177	176	5	9	6.6	18269	PASS