


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD090220\
 Data File : VD066513.D
 Acq On : 02 Sep 2020 09:25
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090220S.M
 Title : SW846 8260
 Last Update : Wed Sep 02 15:08:44 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	21688	PASS
75	95	30	60	59.8	63309	PASS
95	95	100	100	100.0	105946	PASS
96	95	5	9	6.3	6682	PASS
173	174	0.00	2	0.6	633	PASS
174	95	50	100	96.3	102048	PASS
175	174	5	9	9.0	9155	PASS
176	174	95	101	97.7	99733	PASS
177	176	5	9	7.3	7293	PASS