


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD090320\
 Data File : VD066535.D
 Acq On : 03 Sep 2020 13:35
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Title : SW846 8260
 Last Update : Fri Sep 04 01:57:04 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.9	21570	PASS
75	95	30	60	58.7	57840	PASS
95	95	100	100	100.0	98554	PASS
96	95	5	9	6.8	6664	PASS
173	174	0.00	2	1.4	1070	PASS
174	95	50	100	75.9	74845	PASS
175	174	5	9	8.5	6345	PASS
176	174	95	101	95.5	71456	PASS
177	176	5	9	7.4	5283	PASS