

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD090420\
 Data File : VD066561.D
 Acq On : 04 Sep 2020 14:23
 Operator : VA/SY
 Sample : L3923-07
 Misc : 5.01G/5.00ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BP-VPB181-GW-878-880

Quant Time: Sep 05 04:50:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 04 01:57:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	326818	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	598737	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	540908	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	227055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	214337	54.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.42%	
35) Dibromofluoromethane	7.91	113	201069	49.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.56%	
50) Toluene-d8	10.34	98	732942	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
62) 4-Bromofluorobenzene	12.63	95	242838	50.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.46%	
Target Compounds						
16) Acetone	4.16	43	6049	8.314	ug/l #	78
39) Methylcyclohexane	9.34	83	8479	1.219	ug/l	89
40) Benzene	8.34	78	310681	18.205	ug/l	95
67) Ethyl Benzene	11.74	91	224192	10.937	ug/l	100
68) m/p-Xylenes	11.86	106	35152	4.618	ug/l	97
80) 1,3,5-Trimethylbenzene	12.96	105	22932	1.621	ug/l	99
84) 1,2,4-Trimethylbenzene	13.27	105	60103	4.267	ug/l	97
95) Naphthalene	15.41	128	299965	43.981	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD090420\
Data File : VD066561.D
Acq On : 04 Sep 2020 14:23
Operator : VA/SY
Sample : L3923-07
Misc : 5.01G/5.00ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BP-VPB181-GW-878-880

Quant Time: Sep 05 04:50:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 04 01:57:04 2020
Response via : Initial Calibration

