

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
 Data File : VD079773.D
 Acq On : 04 Sep 2024 15:00
 Operator : RP/MD
 Sample : P3813-01RE
 Misc : 4.99G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-04-083024RE

Quant Time: Sep 04 15:53:15 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	14324	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	34355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	34894	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	14060	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	20171	103.279	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	206.560%#
35) Dibromofluoromethane	7.804	113	14866	66.572	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	133.140%
50) Toluene-d8	10.263	98	43587	45.512	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.020%
62) 4-Bromofluorobenzene	12.569	95	12314	45.953	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	91.900%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.805	84	5055	19.123	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
Data File : VD079773.D
Acq On : 04 Sep 2024 15:00
Operator : RP/MD
Sample : P3813-01RE
Misc : 4.99G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-04-083024RE

Quant Time: Sep 04 15:53:15 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

