

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
 Data File : VD079774.D
 Acq On : 04 Sep 2024 15:28
 Operator : RP/MD
 Sample : P3813-03
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TR-06-083024

Quant Time: Sep 04 15:53:24 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	83545	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	174731	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	147952	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	49620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	66860	58.694	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.380%
35) Dibromofluoromethane	7.804	113	62747	55.248	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.500%
50) Toluene-d8	10.269	98	236645	48.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.160%
62) 4-Bromofluorobenzene	12.569	95	49825	36.558	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	73.120%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.799	84	13818	7.522	ug/l	# 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
Data File : VD079774.D
Acq On : 04 Sep 2024 15:28
Operator : RP/MD
Sample : P3813-03
Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TR-06-083024

Quant Time: Sep 04 15:53:24 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

