

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
 Data File : VD079775.D
 Acq On : 04 Sep 2024 15:56
 Operator : RP/MD
 Sample : P3812-01RE
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-083024RE

Quant Time: Sep 04 15:53:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	75121	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	166195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	141491	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	45919	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	72846	71.120	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	142.240%	
35) Dibromofluoromethane	7.804	113	68111	63.051	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	126.100%	
50) Toluene-d8	10.269	98	254222	54.872	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	109.740%	
62) 4-Bromofluorobenzene	12.569	95	55955	43.164	ug/l	0.00
Spiked Amount	50.000	Range 29 - 146	Recovery	=	86.320%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.799	84	18776	12.753	ug/l	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090424\
Data File : VD079775.D
Acq On : 04 Sep 2024 15:56
Operator : RP/MD
Sample : P3812-01RE
Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-083024RE

Quant Time: Sep 04 15:53:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

