

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090524\
 Data File : VD079783.D
 Acq On : 05 Sep 2024 14:11
 Operator : RP/MD
 Sample : VD0905SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01

Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel

Quant Time: Sep 06 01:36:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						09/06/2024
1) Pentafluorobenzene	7.875	168	96695	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	202865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	193941	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	92486	50.000	ug/l	0.00
System Monitoring Compounds						09/06/2024
33) 1,2-Dichloroethane-d4	8.228	65	82622	62.667	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.340%	
35) Dibromofluoromethane	7.804	113	72381	54.892	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.780%	
50) Toluene-d8	10.269	98	322195	56.973	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 113.940%	
62) 4-Bromofluorobenzene	12.569	95	90452	57.163	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 114.320%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.934	85	28996	24.974	ug/l	95
3) Chloromethane	2.146	50	62549	25.860	ug/l	94
4) Vinyl Chloride	2.281	62	74215	25.469	ug/l	100
5) Bromomethane	2.687	94	52249	25.902	ug/l	99
6) Chloroethane	2.834	64	53145	26.902	ug/l	98
7) Trichlorofluoromethane	3.175	101	52486	23.861	ug/l	98
8) Diethyl Ether	3.587	74	17621	25.250	ug/l	74
9) 1,1,2-Trichlorotrifluo...	3.963	101	27368	22.221	ug/l	88
10) Methyl Iodide	4.163	142	29035	20.460	ug/l	90
11) Tert butyl alcohol	5.057	59	8758	134.698	ug/l	99
12) 1,1-Dichloroethene	3.934	96	31329	24.228	ug/l	# 90
13) Acrolein	3.799	56	13612	107.385	ug/l	92
14) Allyl chloride	4.557	41	50313	26.410	ug/l	# 83
15) Acrylonitrile	5.257	53	40974	130.206	ug/l	99
16) Acetone	4.022	43	37497	137.052	ug/l	100
17) Carbon Disulfide	4.275	76	109362	23.129	ug/l	95
18) Methyl Acetate	4.563	43	19267	24.567	ug/l	# 81
19) Methyl tert-butyl Ether	5.322	73	70214	24.811	ug/l	98
20) Methylene Chloride	4.799	84	44972	26.064	ug/l	# 80
21) trans-1,2-Dichloroethene	5.310	96	32921	22.939	ug/l	88
22) Diisopropyl ether	6.216	45	108327	27.019	ug/l	# 88
23) Vinyl Acetate	6.151	43	308017	141.575	ug/l	# 89
24) 1,1-Dichloroethane	6.110	63	66436	25.340	ug/l	97
25) 2-Butanone	7.081	43	48295	132.314	ug/l	96
26) 2,2-Dichloropropane	7.075	77	50126	24.574	ug/l	99
27) cis-1,2-Dichloroethene	7.081	96	38613	23.458	ug/l	83
28) Bromochloromethane	7.428	49	27746	23.811	ug/l	# 56
29) Tetrahydrofuran	7.446	42	30026	132.997	ug/l	# 79
30) Chloroform	7.593	83	63755	24.496	ug/l	97
31) Cyclohexane	7.875	56	52881	23.992	ug/l	# 83
32) 1,1,1-Trichloroethane	7.787	97	52096	24.031	ug/l	95
36) 1,1-Dichloropropene	8.004	75	45238	23.146	ug/l	93
37) Ethyl Acetate	7.169	43	22598	26.165	ug/l	# 94
38) Carbon Tetrachloride	7.987	117	47100	22.597	ug/l	90
39) Methylcyclohexane	9.275	83	49523	20.807	ug/l	# 86
40) Benzene	8.245	78	146316	22.448	ug/l	95

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090524\
 Data File : VD079783.D
 Acq On : 05 Sep 2024 14:11
 Operator : RP/MD
 Sample : VD0905SBS01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01

Quant Time: Sep 06 01:36:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 30 15:57:30 2024
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.404	41	11596	23.463	ug/l	87
42) 1,2-Dichloroethane	8.322	62	38653	24.180	ug/l	92
43) Isopropyl Acetate	8.363	43	39616	24.587	ug/l #	90
44) Trichloroethene	9.028	130	29348	21.098	ug/l	80
45) 1,2-Dichloropropane	9.304	63	35712	22.696	ug/l	97
46) Dibromomethane	9.392	93	18115	21.764	ug/l #	82
47) Bromodichloromethane	9.581	83	47528	22.865	ug/l #	95
48) Methyl methacrylate	9.381	41	18816	24.194	ug/l #	78
49) 1,4-Dioxane	9.387	88	3593	429.666	ug/l #	67
51) 4-Methyl-2-Pentanone	10.157	43	103575	123.062	ug/l	90
52) Toluene	10.328	92	89930	22.813	ug/l	100
53) t-1,3-Dichloropropene	10.551	75	45708	23.325	ug/l	99
54) cis-1,3-Dichloropropene	10.016	75	53654	22.986	ug/l #	93
55) 1,1,2-Trichloroethane	10.728	97	24479	22.481	ug/l #	89
56) Ethyl methacrylate	10.598	69	30605	22.431	ug/l #	76
57) 1,3-Dichloropropane	10.875	76	42415	21.608	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.869	63	62567	111.250	ug/l #	85
59) 2-Hexanone	10.922	43	73574	124.230	ug/l	86
60) Dibromochloromethane	11.069	129	29004	21.655	ug/l	98
61) 1,2-Dibromoethane	11.175	107	20879	21.879	ug/l	96
64) Tetrachloroethene	10.810	164	22756	19.975	ug/l	96
65) Chlorobenzene	11.604	112	93099	21.635	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.675	131	29656	20.980	ug/l	99
67) Ethyl Benzene	11.681	91	159994	21.091	ug/l	94
68) m/p-Xylenes	11.786	106	124065	42.605	ug/l	88
69) o-Xylene	12.116	106	56287	20.739	ug/l	84
70) Styrene	12.134	104	105907	22.356	ug/l	95
71) Bromoform	12.292	173	15111	21.317	ug/l #	100
73) Isopropylbenzene	12.416	105	144806	21.999	ug/l	97
74) N-amyl acetate	12.228	43	40935	26.165	ug/l #	84
75) 1,1,2,2-Tetrachloroethane	12.669	83	27853	22.044	ug/l	95
76) 1,2,3-Trichloropropane	12.722	75	18265m	23.106	ug/l	
77) Bromobenzene	12.698	156	31199	21.297	ug/l	72
78) n-propylbenzene	12.757	91	188957	23.231	ug/l	93
79) 2-Chlorotoluene	12.845	91	101252	22.345	ug/l	93
80) 1,3,5-Trimethylbenzene	12.898	105	122994	22.476	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.463	75	11071	23.360	ug/l	93
82) 4-Chlorotoluene	12.939	91	110933	23.505	ug/l	93
83) tert-Butylbenzene	13.163	119	101826	22.392	ug/l	92
84) 1,2,4-Trimethylbenzene	13.210	105	128603	22.781	ug/l	95
85) sec-Butylbenzene	13.339	105	156984	22.579	ug/l	97
86) p-Isopropyltoluene	13.457	119	128200	21.578	ug/l	96
87) 1,3-Dichlorobenzene	13.457	146	66741	21.221	ug/l	96
88) 1,4-Dichlorobenzene	13.533	146	65826	20.934	ug/l	93
89) n-Butylbenzene	13.780	91	125047	22.450	ug/l	98
90) Hexachloroethane	14.045	117	26253	22.331	ug/l	80
91) 1,2-Dichlorobenzene	13.828	146	57714	21.429	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.445	75	4499	24.388	ug/l	62
93) 1,2,4-Trichlorobenzene	15.092	180	29765	20.121	ug/l	96
94) Hexachlorobutadiene	15.198	225	15080	21.199	ug/l	95
95) Naphthalene	15.322	128	67350	22.436	ug/l #	96
96) 1,2,3-Trichlorobenzene	15.516	180	27055	20.931	ug/l	97

09/06/2024
 Supervised By :Mahesh
 Dadoda

09/06/2024

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090524\
Data File : VD079783.D
Acq On : 05 Sep 2024 14:11
Operator : RP/MD
Sample : VD0905SBS01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Quant Time: Sep 06 01:36:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D083024S.M
Quant Title : SW846 8260
QLast Update : Fri Aug 30 15:57:30 2024
Response via : Initial Calibration

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Manual Integrations
APPROVED


Reviewed By :Romaben
Patel

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#) = qualifier out of range (m) = manual integration (+) = signals summed						

09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.875 min Scan# 1069

Delta R.T. -0.000 min

Lab File: VD079783.D


Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :

VD0905SBS01



Tgt Ion:168 Resp: 9669

Ion Ratio Lower Upper


168 100

99 66.5 41.1 61.7

Manual Integrations

APPROVED


Reviewed By :Romaben
Patel



09/06/2024

Supervised By :Mahesh
Dadoda

09/06/2024



#2

Dichlorodifluoromethane

Concen: 24.974 ug/l

RT: 1.934 min Scan# 59

Delta R.T. 0.000 min

Lab File: VD079783.D




Acq: 05 Sep 2024 14:11


Tgt Ion: 85 Resp: 28996

Ion Ratio Lower Upper

85 100


87 35.0 16.0 48.0





#3
Chloromethane
Concen: 25.860 ug/l
RT: 2.146 min Scan# 95
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 50 Resp: 62549
Ion Ratio Lower Upper
50 100
52 31.5 28.2 42.2


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#4
Vinyl Chloride
Concen: 25.469 ug/l
RT: 2.281 min Scan# 118
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 62 Resp: 74215
Ion Ratio Lower Upper
62 100
64 32.7 26.3 39.5





#5
Bromomethane
Concen: 25.902 ug/l
RT: 2.687 min Scan# 188
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 94 Resp: 52249
Ion Ratio Lower Upper
94 100
96 94.7 76.9 115.3


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#6
Chloroethane
Concen: 26.902 ug/l
RT: 2.834 min Scan# 212
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 64 Resp: 53145
Ion Ratio Lower Upper
64 100
66 29.5 24.5 36.7





#7
Trichlorofluoromethane
Concen: 23.861 ug/l
RT: 3.175 min Scan# 21
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:101 Resp: 52480
Ion Ratio Lower Upper
101 100
103 65.0 53.2 79.8


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#8
Diethyl Ether
Concen: 25.250 ug/l
RT: 3.587 min Scan# 340
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion: 74 Resp: 17621
Ion Ratio Lower Upper
74 100
45 98.2 37.9 113.6







#9
1,1,2-Trichlorotrifluoroethane
Concen: 22.221 ug/l
RT: 3.963 min Scan# 404
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion:101 Resp: 2736
Ion Ratio Lower Upper
101 100
85 48.8 33.5 50.3
151 65.1 60.3 90.5



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#10
Methyl Iodide
Concen: 20.460 ug/l
RT: 4.163 min Scan# 438
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:142 Resp: 29035
Ion Ratio Lower Upper
142 100
127 50.4 34.2 51.4
141 16.0 11.7 17.5





#11
Tert butyl alcohol
Concen: 134.698 ug/l
RT: 5.057 min Scan# 590
Delta R.T. 0.005 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 59 Resp: 8758
Ion Ratio Lower Upper
59 100
57 4.9 3.6 5.4


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#12
1,1-Dichloroethene
Concen: 24.228 ug/l
RT: 3.934 min Scan# 399
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 96 Resp: 31329
Ion Ratio Lower Upper
96 100
61 142.8 107.3 160.9
98 51.2 51.7 77.5#





#13
Acrolein
Concen: 107.385 ug/l
RT: 3.799 min Scan# 31
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 56 Resp: 1361
Ion Ratio Lower Upper
56 100
55 75.6 55.3 82.9


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#14
Allyl chloride
Concen: 26.410 ug/l
RT: 4.557 min Scan# 505
Delta R.T. -0.007 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 41 Resp: 50313
Ion Ratio Lower Upper
41 100
39 62.2 58.0 87.0
76 40.8 45.9 68.9#





#15
Acrylonitrile
Concen: 130.206 ug/l
RT: 5.257 min Scan# 61
Delta R.T. 0.005 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 53 Resp: 40974
Ion Ratio Lower Upper
53 100
52 85.4 67.7 101.5
51 37.2 28.6 42.8


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#16
Acetone
Concen: 137.052 ug/l
RT: 4.022 min Scan# 414
Delta R.T. 0.005 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Tgt Ion: 43 Resp: 37497
Ion Ratio Lower Upper
43 100
58 37.8 30.4 45.6







#17
Carbon Disulfide
Concen: 23.129 ug/l
RT: 4.275 min Scan# 457
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion: 76 Resp: 109362
Ion Ratio Lower Upper
76 100
78 6.9 6.9 10.3



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#18
Methyl Acetate
Concen: 24.567 ug/l
RT: 4.563 min Scan# 506
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 43 Resp: 19267
Ion Ratio Lower Upper
43 100
74 24.2 28.1 42.1#





#19
Methyl tert-butyl Ether
Concen: 24.811 ug/l
RT: 5.322 min Scan# 61
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 73 Resp: 70214
Ion Ratio Lower Upper
73 100
57 22.1 18.3 27.5


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#20
Methylene Chloride
Concen: 26.064 ug/l
RT: 4.799 min Scan# 546
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 84 Resp: 44972
Ion Ratio Lower Upper
84 100
49 123.9 75.6 113.4#
51 40.1 24.2 36.2#
86 57.0 49.2 73.8





#21
trans-1,2-Dichloroethene
Concen: 22.939 ug/l
RT: 5.310 min Scan# 611
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 96 Resp: 3292
Ion Ratio Lower Upper
96 100
61 125.7 84.5 126.7
98 62.5 50.5 75.7


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#22
Diisopropyl ether
Concen: 27.019 ug/l
RT: 6.216 min Scan# 787
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 45 Resp: 108327
Ion Ratio Lower Upper
45 100
43 55.2 41.7 62.5
87 22.3 30.2 45.2#
59 12.7 11.0 16.6





#23
 Vinyl Acetate
 Concen: 141.575 ug/l
 RT: 6.151 min Scan# 776
 Delta R.T. -0.001 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11



Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01



Tgt Ion: 43 Resp: 30801
 Ion Ratio Lower Upper
 43 100
 86 10.7 12.3 18.5


Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel






09/06/2024
 Supervised By :Mahesh
 Dadoda


09/06/2024



#24
 1,1-Dichloroethane
 Concen: 25.340 ug/l
 RT: 6.110 min Scan# 769
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11


Tgt Ion: 63 Resp: 66436
 Ion Ratio Lower Upper
 63 100
 98 8.2 4.7 14.0
 100 4.5 2.9 8.6





#25
2-Butanone
Concen: 132.314 ug/l
RT: 7.081 min Scan# 91
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 43 Resp: 4829
Ion Ratio Lower Upper
43 100
72 30.5 26.4 39.6


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#26
2,2-Dichloropropane
Concen: 24.574 ug/l
RT: 7.075 min Scan# 933
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 77 Resp: 50126
Ion Ratio Lower Upper
77 100
97 24.1 12.4 37.2





#27
 cis-1,2-Dichloroethene
 Concen: 23.458 ug/l
 RT: 7.081 min Scan# 91
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11



Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01



Tgt Ion: 96 Resp: 3861
 Ion Ratio Lower Upper
 96 100
 61 143.1 0.0 229.0
 98 64.6 0.0 127.2


Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel






09/06/2024
 Supervised By :Mahesh
 Dadoda


09/06/2024



#28
 Bromochloromethane
 Concen: 23.811 ug/l
 RT: 7.428 min Scan# 993
 Delta R.T. 0.006 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11

Tgt Ion: 49 Resp: 27746
 Ion Ratio Lower Upper
 49 100
 129 0.9 0.0 7.0
 130 61.5 87.7 131.5#





#29

Tetrahydrofuran

Concen: 132.997 ug/l

RT: 7.446 min Scan# 996

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D



ClientSampleId :

VD0905SBS01

Tgt Ion:	42	Resp:	30020
Ion Ratio	Lower	Upper	
42	100		
72	44.3	49.4	74.0
71	42.0	45.0	67.6



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel



09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#30

Chloroform

Concen: 24.496 ug/l




RT: 7.593 min Scan# 1021


Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Tgt Ion:	83	Resp:	63755
Ion Ratio	Lower	Upper	
83	100		
85	62.1	51.8	77.6





#31

Cyclohexane

Concen: 23.992 ug/l

RT: 7.875 min Scan# 1069

Delta R.T. -0.000 min

Lab File: VD079783.D


Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :


VD0905SBS01



Tgt Ion:	56	Resp:	5288
Ion Ratio	Lower	Upper	
56	100		
69	35.3	28.0	42.0
84	86.0	87.4	131.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel



09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#32

1,1,1-Trichloroethane

Concen: 24.031 ug/l




RT: 7.787 min Scan# 1054


Delta R.T. -0.006 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Tgt Ion:	97	Resp:	52096
Ion Ratio	Lower	Upper	
97	100		
99	63.9	51.3	76.9
61	46.5	31.4	47.2





#33

1,2-Dichloroethane-d4

Concen: 62.667 ug/l

RT: 8.228 min Scan# 1129

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :

VD0905SBS01

Tgt Ion: 65 Resp: 8262

Ion Ratio Lower Upper



65 100

67 53.8 0.0 111.0

Manual Integrations

APPROVED



Reviewed By :Romaben
Patel



09/06/2024

Supervised By :Mahesh
Dadoda

09/06/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.775 min Scan# 1222

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11




Tgt Ion:114 Resp: 202865


Ion Ratio Lower Upper

114 100

63 21.2 0.0 34.0

88 15.6 0.0 27.2





#35

Dibromofluoromethane

Concen: 54.892 ug/l

RT: 7.804 min Scan# 1057

Delta R.T. -0.001 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :



VD0905SBS01

Tgt Ion: 113 Resp: 7238

Ion	Ratio	Lower	Upper
113	100		
111	102.1	84.0	126.0
192	15.0	15.4	23.0



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel



09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#36

1,1-Dichloropropene

Concen: 23.146 ug/l

RT: 8.004 min Scan# 1091




Delta R.T. -0.000 min


Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Tgt Ion: 75 Resp: 45238


Ion	Ratio	Lower	Upper
75	100		
110	32.1	19.6	58.7
77	31.6	26.1	39.1





#37
Ethyl Acetate
Concen: 26.165 ug/l
RT: 7.169 min Scan# 949
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 43 Resp: 2259
Ion Ratio Lower Upper
43 100
61 15.3 13.9 20.9
70 9.6 9.8 14.6


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#38
Carbon Tetrachloride
Concen: 22.597 ug/l
RT: 7.987 min Scan# 1088
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion:117 Resp: 47100
Ion Ratio Lower Upper
117 100
119 84.4 75.4 113.0
121 25.3 25.0 37.4





#39

Methylcyclohexane

Concen: 20.807 ug/l

RT: 9.275 min Scan# 11307

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :



VD0905SBS01

Tgt Ion: 83 Resp: 4952

Ion	Ratio	Lower	Upper
83	100		
55	80.8	50.1	75.1
98	51.1	41.9	62.9



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel



09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#40

Benzene

Concen: 22.448 ug/l

RT: 8.245 min Scan# 1132




Delta R.T. -0.000 min


Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Tgt Ion: 78 Resp: 146316


Ion	Ratio	Lower	Upper
78	100		
77	24.8	18.1	27.1





#41
Methacrylonitrile
Concen: 23.463 ug/l
RT: 7.404 min Scan# 91
Delta R.T. 0.005 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:	41	Resp:	11590
Ion	Ratio	Lower	Upper
41	100		
39	56.9	41.4	62.0
67	87.8	85.8	128.6
52	35.4	30.4	45.6


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#42
1,2-Dichloroethane
Concen: 24.180 ug/l
RT: 8.322 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:	62	Resp:	38653
Ion	Ratio	Lower	Upper
62	100		
98	9.7	0.0	25.8





#43
Isopropyl Acetate
Concen: 24.587 ug/l
RT: 8.363 min Scan# 1152
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 43 Resp: 39610
Ion Ratio Lower Upper
43 100
61 31.9 29.3 43.9
87 12.4 14.9 22.3


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#44
Trichloroethene
Concen: 21.098 ug/l
RT: 9.028 min Scan# 1265
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:130 Resp: 29348
Ion Ratio Lower Upper
130 100
95 107.6 0.0 177.8





#45
1,2-Dichloropropane
Concen: 22.696 ug/l
RT: 9.304 min Scan# 1312
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 63 Resp: 35712
Ion Ratio Lower Upper
63 100
65 31.3 23.8 35.8


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#46
Dibromomethane
Concen: 21.764 ug/l
RT: 9.392 min Scan# 1327
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 93 Resp: 18115
Ion Ratio Lower Upper
93 100
95 84.2 65.6 98.4
174 72.4 82.3 123.5#





#47
Bromodichloromethane
Concen: 22.865 ug/l
RT: 9.581 min Scan# 1359
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 83 Resp: 47528
Ion Ratio Lower Upper
83 100
85 66.2 50.1 75.1
127 6.5 6.8 10.2


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#48
Methyl methacrylate
Concen: 24.194 ug/l
RT: 9.381 min Scan# 1325
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Tgt Ion: 41 Resp: 18816
Ion Ratio Lower Upper
41 100
69 90.4 96.7 145.1#
39 47.3 43.8 65.6





#49

1,4-Dioxane

Concen: 429.666 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :



VD0905SBS01

Tgt Ion: 88 Resp: 359

Ion	Ratio	Lower	Upper
88	100		
43	39.7	21.5	32.3
58	79.3	42.2	63.4



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel



09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#50

Toluene-d8

Concen: 56.973 ug/l

RT: 10.269 min Scan# 1476




Delta R.T. -0.000 min


Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Tgt Ion: 98 Resp: 322195

Ion	Ratio	Lower	Upper
98	100		
100	61.6	51.6	77.4







#51
4-Methyl-2-Pentanone
Concen: 123.062 ug/l
RT: 10.157 min Scan# 1457
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion: 43 Resp: 103579
Ion Ratio Lower Upper
43 100
58 39.1 36.6 55.0



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#52
Toluene
Concen: 22.813 ug/l
RT: 10.328 min Scan# 1486
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 92 Resp: 89930
Ion Ratio Lower Upper
92 100
91 176.6 141.7 212.5





#53
 t-1,3-Dichloropropene
 Concen: 23.325 ug/l
 RT: 10.551 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11



Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01



Tgt Ion: 75 Resp: 45708
 Ion Ratio Lower Upper
 75 100
 77 31.8 25.0 37.6


Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel






09/06/2024
 Supervised By :Mahesh
 Dadoda


09/06/2024



#54
 cis-1,3-Dichloropropene
 Concen: 22.986 ug/l
 RT: 10.016 min Scan# 1433
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11


Tgt Ion: 75 Resp: 53654
 Ion Ratio Lower Upper
 75 100
 77 30.2 24.4 36.6
 39 37.6 24.2 36.2#





#55
1,1,2-Trichloroethane
Concen: 22.481 ug/l
RT: 10.728 min Scan# 1554
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument : MSVOA_D
ClientSampleId : VD0905SBS01



Tgt Ion:	97	Resp:	24479
Ion Ratio	100	Lower	Upper
97	100		
83	83.4	65.8	98.8
85	43.5	43.8	65.8
99	53.4	55.4	83.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#56
Ethyl methacrylate
Concen: 22.431 ug/l
RT: 10.598 min Scan# 1532
Delta R.T. 0.005 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion:	69	Resp:	30605
Ion Ratio	100	Lower	Upper
69	100		
41	67.2	39.8	59.6
39	36.2	20.2	30.4





#57

1,3-Dichloropropane

Concen: 21.608 ug/l

RT: 10.875 min Scan# 1579

Delta R.T. -0.000 min

Lab File: VD079783.D

Acq: 05 Sep 2024 14:11

Instrument :

MSVOA_D

ClientSampleId :

VD0905SBS01

Tgt Ion: 76 Resp: 4241

Ion Ratio Lower Upper



76 100

78 33.2 26.2 39.2

Manual Integrations

APPROVED



Reviewed By :Romaben
Patel



09/06/2024

Supervised By :Mahesh
Dadoda

09/06/2024



#58

2-Chloroethyl Vinyl ether

Concen: 111.250 ug/l

RT: 9.869 min Scan# 1408

Delta R.T. -0.000 min

Lab File: VD079783.D




Acq: 05 Sep 2024 14:11


Tgt Ion: 63 Resp: 62567

Ion Ratio Lower Upper

63 100


106 24.6 26.4 39.6





#59
2-Hexanone
Concen: 124.230 ug/l
RT: 10.922 min Scan# 11
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 43 Resp: 73574
Ion Ratio Lower Upper
43 100
58 51.8 31.1 93.5


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#60
Dibromochloromethane
Concen: 21.655 ug/l
RT: 11.069 min Scan# 1612
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:129 Resp: 29004
Ion Ratio Lower Upper
129 100
127 77.8 38.0 113.9





#61
1,2-Dibromoethane
Concen: 21.879 ug/l
RT: 11.175 min Scan# 1630
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:107 Resp: 20879
Ion Ratio Lower Upper
107 100
109 94.0 72.0 108.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#62
4-Bromofluorobenzene
Concen: 57.163 ug/l
RT: 12.569 min Scan# 1867
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 95 Resp: 90452
Ion Ratio Lower Upper
95 100
174 75.2 0.0 195.4
176 70.2 0.0 192.6





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.581 min Scan# 11
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:117 Resp: 19394
Ion Ratio Lower Upper
117 100
82 56.6 39.8 59.6
119 29.0 25.7 38.5


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#64
Tetrachloroethene
Concen: 19.975 ug/l
RT: 10.810 min Scan# 1568
Delta R.T. 0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:164 Resp: 22756
Ion Ratio Lower Upper
164 100
166 127.6 108.6 163.0
129 93.0 74.6 112.0
131 89.9 68.6 103.0





#65
Chlorobenzene
Concen: 21.635 ug/l
RT: 11.604 min Scan# 1703
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:112 Resp: 93099
Ion Ratio Lower Upper
112 100
114 32.8 26.0 39.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 20.980 ug/l
RT: 11.675 min Scan# 1715
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion:131 Resp: 29656
Ion Ratio Lower Upper
131 100
133 97.5 49.2 147.6
119 65.6 33.4 100.2







#67
Ethyl Benzene
Concen: 21.091 ug/l
RT: 11.681 min Scan# 1716
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion: 91 Resp: 159994
Ion Ratio Lower Upper
91 100
106 28.7 25.8 38.6



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#68
m/p-Xylenes
Concen: 42.605 ug/l
RT: 11.786 min Scan# 1734
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:106 Resp: 124065
Ion Ratio Lower Upper
106 100
91 202.8 148.5 222.7





#69
o-Xylene
Concen: 20.739 ug/l
RT: 12.116 min Scan# 1790
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:106 Resp: 5628
Ion Ratio Lower Upper
106 100
91 216.6 96.7 290.1


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#70
Styrene
Concen: 22.356 ug/l
RT: 12.134 min Scan# 1793
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:104 Resp: 105907
Ion Ratio Lower Upper
104 100
78 47.1 33.0 49.6
103 53.6 44.2 66.4





#71
Bromoform
Concen: 21.317 ug/l
RT: 12.292 min Scan# 1821
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:173 Resp: 1511
Ion Ratio Lower Upper
173 100
175 47.2 23.7 71.1
254 0.0 0.0 0.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.516 min Scan# 2028
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:152 Resp: 92486
Ion Ratio Lower Upper
152 100
115 63.1 26.5 79.5
150 167.4 0.0 341.6





#73
Isopropylbenzene
Concen: 21.999 ug/l
RT: 12.416 min Scan# 1841
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:105 Resp: 144800
Ion Ratio Lower Upper
105 100
120 25.6 13.6 40.7


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#74
N-ethyl acetate
Concen: 26.165 ug/l
RT: 12.228 min Scan# 1809
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 43 Resp: 40935
Ion Ratio Lower Upper
43 100
70 40.9 44.3 66.5#
55 23.3 23.2 34.8
61 22.8 24.2 36.4#





#75
1,1,2,2-Tetrachloroethane
Concen: 22.044 ug/l
RT: 12.669 min Scan# 1884
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 83 Resp: 2785
Ion Ratio Lower Upper
83 100
131 9.3 5.5 16.5
85 63.8 33.8 101.3


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#76
1,2,3-Trichloropropane
Concen: 23.106 ug/l m
RT: 12.722 min Scan# 1893
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Tgt Ion: 75 Resp: 18265
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0







#77
Bromobenzene
Concen: 21.297 ug/l
RT: 12.698 min Scan# 1889
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion:156 Resp: 31199
Ion Ratio Lower Upper
156 100
77 208.4 77.3 231.9
158 91.8 49.6 149.0



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#78
n-propylbenzene
Concen: 23.231 ug/l
RT: 12.757 min Scan# 1899
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 91 Resp: 188957
Ion Ratio Lower Upper
91 100
120 21.0 12.3 36.8





#79
2-Chlorotoluene
Concen: 22.345 ug/l
RT: 12.845 min Scan# 1914
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 91 Resp: 101252
Ion Ratio Lower Upper
91 100
126 34.1 19.3 57.8


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#80
1,3,5-Trimethylbenzene
Concen: 22.476 ug/l
RT: 12.898 min Scan# 1923
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:105 Resp: 122994
Ion Ratio Lower Upper
105 100
120 48.3 25.7 77.1





#81
trans-1,4-Dichloro-2-butene
Concen: 23.360 ug/l
RT: 12.463 min Scan# 1107
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion: 75 Resp: 1107
Ion Ratio Lower Upper
75 100
53 84.4 61.4 92.2
89 45.9 35.4 53.0


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#82
4-Chlorotoluene
Concen: 23.505 ug/l
RT: 12.939 min Scan# 1930
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion: 91 Resp: 110933
Ion Ratio Lower Upper
91 100
126 33.8 18.9 56.9





#83
tert-Butylbenzene
Concen: 22.392 ug/l
RT: 13.163 min Scan# 1968
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:119 Resp: 101820
Ion Ratio Lower Upper
119 100
91 64.6 28.2 84.7
134 26.1 13.4 40.1


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#84
1,2,4-Trimethylbenzene
Concen: 22.781 ug/l
RT: 13.210 min Scan# 1976
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion:105 Resp: 128603
Ion Ratio Lower Upper
105 100
120 44.0 23.6 71.0





#85
 sec-Butylbenzene
 Concen: 22.579 ug/l
 RT: 13.339 min Scan# 1998
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11



Instrument :
 MSVOA_D
 ClientSampleId :
 VD0905SBS01



Tgt Ion:105 Resp: 156984
 Ion Ratio Lower Upper
 105 100
 134 19.3 10.3 30.8


Manual Integrations
 APPROVED

Reviewed By :Romaben
 Patel




09/06/2024
 Supervised By :Mahesh
 Dadoda



09/06/2024




#86
 p-Isopropyltoluene
 Concen: 21.578 ug/l
 RT: 13.457 min Scan# 2018
 Delta R.T. -0.000 min
 Lab File: VD079783.D
 Acq: 05 Sep 2024 14:11




Tgt Ion:119 Resp: 128200
 Ion Ratio Lower Upper
 119 100
 134 26.0 13.7 41.1
 91 24.4 10.7 32.1





#87
1,3-Dichlorobenzene
Concen: 21.221 ug/l
RT: 13.457 min Scan# 2018
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:146 Resp: 6674
Ion Ratio Lower Upper
146 100
111 42.5 18.5 55.5
148 62.1 30.9 92.8


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#88
1,4-Dichlorobenzene
Concen: 20.934 ug/l
RT: 13.533 min Scan# 2031
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion:146 Resp: 65826
Ion Ratio Lower Upper
146 100
111 41.9 17.5 52.6
148 66.3 31.4 94.2







#89
n-Butylbenzene
Concen: 22.450 ug/l
RT: 13.780 min Scan# 2074
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11

Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01

Tgt Ion: 91 Resp: 12504
Ion Ratio Lower Upper
91 100
92 51.4 25.7 77.0
134 23.8 13.6 40.8



Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#90
Hexachloroethane
Concen: 22.331 ug/l
RT: 14.045 min Scan# 2118
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:117 Resp: 26253
Ion Ratio Lower Upper
117 100
201 55.8 36.3 108.9





#91
1,2-Dichlorobenzene
Concen: 21.429 ug/l
RT: 13.828 min Scan# 2081
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:146 Resp: 57714
Ion Ratio Lower Upper
146 100
111 43.1 19.1 57.1
148 64.3 31.1 93.5


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda



09/06/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 24.388 ug/l
RT: 14.445 min Scan# 2186
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11




Tgt Ion: 75 Resp: 4499
Ion Ratio Lower Upper
75 100
155 60.1 47.7 143.1
157 82.2 63.4 190.2





#93
1,2,4-Trichlorobenzene
Concen: 20.121 ug/l
RT: 15.092 min Scan# 21
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:180 Resp: 2976
Ion Ratio Lower Upper
180 100
182 94.8 46.9 140.7
145 34.9 14.0 41.9


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel






09/06/2024
Supervised By :Mahesh
Dadoda


09/06/2024



#94
Hexachlorobutadiene
Concen: 21.199 ug/l
RT: 15.198 min Scan# 2314
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11


Tgt Ion:225 Resp: 15080
Ion Ratio Lower Upper
225 100
223 63.1 33.3 99.9
227 63.0 33.6 100.6





#95
Naphthalene
Concen: 22.436 ug/l
RT: 15.322 min Scan# 21
Delta R.T. -0.006 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Instrument :
MSVOA_D
ClientSampleId :
VD0905SBS01



Tgt Ion:128 Resp: 67350
Ion Ratio Lower Upper
128 100
127 12.6 10.2 15.4
129 13.3 8.5 12.7


Manual Integrations
APPROVED

Reviewed By :Romaben
Patel




09/06/2024
Supervised By :Mahesh
Dadoda

09/06/2024



#96
1,2,3-Trichlorobenzene
Concen: 20.931 ug/l
RT: 15.516 min Scan# 2368
Delta R.T. -0.000 min
Lab File: VD079783.D
Acq: 05 Sep 2024 14:11



Tgt Ion:180 Resp: 27055
Ion Ratio Lower Upper
180 100
182 93.2 47.8 143.3
145 35.6 16.1 48.2

