

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD090618\
 Data File : VD059980.D
 Acq On : 6 Sep 2018 12:24
 Operator : JC/SY
 Sample : VD0906SBL01
 Misc : 5.00g/5ml/MSVOA D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0906SBL01

Quant Time: Sep 07 04:05:09 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 23 03:03:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.38	168	1280213	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.43	114	1718222	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.27	117	1404285	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	13.37	152	746858	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.77	65	517368	47.95	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	95.90%	
35) Dibromofluoromethane	6.30	113	682476	50.66	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	9.44	98	1909409	51.07	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	12.42	95	757059	49.86	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	99.72%	
Target Compounds						
20) Methylene Chloride	3.69	84	158495	9.03	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD090618\
Data File : VD059980.D
Acq On : 6 Sep 2018 12:24
Operator : JC/SY
Sample : VD0906SBL01
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0906SBL01

Quant Time: Sep 07 04:05:09 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D082218S.M
Quant Title : SW846 8260
QLast Update : Thu Aug 23 03:03:08 2018
Response via : Initial Calibration

