


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD090920\
Data File : VD066573.D
Acq On : 09 Sep 2020 09:34
Operator : VA/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
Title : SW846 8260
Last Update : Fri Sep 04 01:57:04 2020



AutoFind: Scans 1820, 1821, 1822; Background Corrected with Scan 1812

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.3	51842	PASS
75	95	30	60	55.8	129504	PASS
95	95	100	100	100.0	232104	PASS
96	95	5	9	7.0	16320	PASS
173	174	0.00	2	1.0	1767	PASS
174	95	50	100	76.4	177280	PASS
175	174	5	9	7.5	13288	PASS
176	174	95	101	96.1	170453	PASS
177	176	5	9	7.3	12430	PASS