

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090922\
 Data File : VD074245.D
 Acq On : 09 Sep 2022 16:55
 Operator : VA/SY
 Sample : VSTDICCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022

Quant Time: Sep 10 02:46:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 02:44:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	127122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.769	114	206717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.569	117	186309	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.498	152	88587	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	63888	48.941	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	97.880%		
35) Dibromofluoromethane	7.799	113	67657	49.087	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.180%		
50) Toluene-d8	10.263	98	262616	57.175	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery =	114.340%		
62) 4-Bromofluorobenzene	12.557	95	83434	48.226	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery =	96.460%		
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.928	85	46670	44.557	ug/l	100
3) Chloromethane	2.140	50	40669	53.209	ug/l	100
4) Vinyl Chloride	2.275	62	37436	48.149	ug/l	100
5) Bromomethane	2.675	94	19227	39.412	ug/l	100
6) Chloroethane	2.822	64	21014	42.601	ug/l	100
7) Trichlorofluoromethane	3.158	101	102500	45.475	ug/l	100
8) Diethyl Ether	3.587	74	34120	54.431	ug/l	100
9) 1,1,2-Trichlorotrifluo...	3.952	101	66745	48.366	ug/l	100
10) Methyl Iodide	4.152	142	84871	56.077	ug/l	100
11) Tert butyl alcohol	5.063	59	23730	282.143	ug/l	100
12) 1,1-Dichloroethene	3.928	96	64621	52.010	ug/l	100
13) Acrolein	3.787	56	27579	536.906	ug/l	100
14) Allyl chloride	4.552	41	100394	51.753	ug/l	100
15) Acrylonitrile	5.252	53	82920	281.624	ug/l	100
16) Acetone	4.016	43	73902	294.006	ug/l	100
17) Carbon Disulfide	4.258	76	182282	51.254	ug/l	100
18) Methyl Acetate	4.558	43	40246	50.907	ug/l	100
19) Methyl tert-butyl Ether	5.316	73	161772	53.553	ug/l	100
20) Methylene Chloride	4.793	84	78281	40.719	ug/l	100
21) trans-1,2-Dichloroethene	5.305	96	73308	50.870	ug/l	100
22) Diisopropyl ether	6.210	45	221697	54.736	ug/l	100
23) Vinyl Acetate	6.152	43	550952	269.883	ug/l	100
24) 1,1-Dichloroethane	6.105	63	135216	51.124	ug/l	100
25) 2-Butanone	7.075	43	109117	279.615	ug/l	100
26) 2,2-Dichloropropane	7.069	77	116501	49.208	ug/l	100
27) cis-1,2-Dichloroethene	7.075	96	84292	51.530	ug/l	100
28) Bromochloromethane	7.422	49	45476	55.544	ug/l	100
29) Tetrahydrofuran	7.440	42	66459	282.633	ug/l	100
30) Chloroform	7.587	83	132387	48.103	ug/l	100
31) Cyclohexane	7.869	56	117478	51.471	ug/l	100
32) 1,1,1-Trichloroethane	7.787	97	117026	46.100	ug/l	100
36) 1,1-Dichloropropene	8.004	75	103980	50.717	ug/l	100
37) Ethyl Acetate	7.163	43	45797	55.100	ug/l	100
38) Carbon Tetrachloride	7.981	117	97839	45.193	ug/l	100
39) Methylcyclohexane	9.263	83	127970	54.504	ug/l	100
40) Benzene	8.240	78	293482	50.507	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090922\
 Data File : VD074245.D
 Acq On : 09 Sep 2022 16:55
 Operator : VA/SY
 Sample : VSTDICCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022

Quant Time: Sep 10 02:46:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 02:44:06 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.399	41	29311	57.864	ug/l	100
42) 1,2-Dichloroethane	8.316	62	77202	45.647	ug/l	100
43) Isopropyl Acetate	8.352	43	90110	53.816	ug/l	100
44) Trichloroethene	9.022	130	81326	48.428	ug/l	100
45) 1,2-Dichloropropane	9.299	63	75938	51.800	ug/l	100
46) Dibromomethane	9.387	93	40098	49.616	ug/l	100
47) Bromodichloromethane	9.575	83	100932	48.138	ug/l	100
48) Methyl methacrylate	9.375	41	42641	55.689	ug/l	100
49) 1,4-Dioxane	9.375	88	10548	1118.293	ug/l	100
51) 4-Methyl-2-Pentanone	10.151	43	233313	279.250	ug/l	100
52) Toluene	10.328	92	189516	49.873	ug/l	100
53) t-1,3-Dichloropropene	10.546	75	97521	51.012	ug/l	100
54) cis-1,3-Dichloropropene	10.010	75	116661	51.409	ug/l	100
55) 1,1,2-Trichloroethane	10.722	97	56474	50.529	ug/l	100
56) Ethyl methacrylate	10.587	69	76791	59.253	ug/l	100
57) 1,3-Dichloropropane	10.869	76	98013	52.433	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.863	63	60136	163.069	ug/l	100
59) 2-Hexanone	10.910	43	163212	289.845	ug/l	100
60) Dibromochloromethane	11.063	129	67886	47.601	ug/l	100
61) 1,2-Dibromoethane	11.169	107	53308	49.466	ug/l	100
64) Tetrachloroethene	10.798	164	65942	49.167	ug/l	100
65) Chlorobenzene	11.598	112	201031	49.897	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.669	131	72307	47.988	ug/l	100
67) Ethyl Benzene	11.669	91	373141	53.827	ug/l	100
68) m/p-Xylenes	11.781	106	283947	104.023	ug/l	100
69) o-Xylene	12.110	106	135088	53.307	ug/l	100
70) Styrene	12.122	104	227820	52.602	ug/l	100
71) Bromoform	12.287	173	41365	50.020	ug/l #	100
73) Isopropylbenzene	12.404	105	359181	55.774	ug/l	100
74) N-amyl acetate	12.216	43	87417	60.708	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.657	83	66485	56.676	ug/l	100
76) 1,2,3-Trichloropropane	12.710	75	48155m	54.106	ug/l	100
77) Bromobenzene	12.687	156	78038	51.304	ug/l	100
78) n-propylbenzene	12.745	91	436542	55.412	ug/l	100
79) 2-Chlorotoluene	12.834	91	238222	52.703	ug/l	100
80) 1,3,5-Trimethylbenzene	12.887	105	292246	53.293	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.451	75	18082	53.507	ug/l #	100
82) 4-Chlorotoluene	12.928	91	240486	50.656	ug/l	100
83) tert-Butylbenzene	13.151	119	258113	55.223	ug/l	100
84) 1,2,4-Trimethylbenzene	13.192	105	288287	53.887	ug/l	100
85) sec-Butylbenzene	13.328	105	390249	55.364	ug/l	100
86) p-Isopropyltoluene	13.440	119	322343	55.670	ug/l	100
87) 1,3-Dichlorobenzene	13.440	146	158442	51.620	ug/l	98
88) 1,4-Dichlorobenzene	13.522	146	156276	50.531	ug/l	100
89) n-Butylbenzene	13.769	91	313631	56.623	ug/l	100
90) Hexachloroethane	14.034	117	58906	50.262	ug/l	100
91) 1,2-Dichlorobenzene	13.810	146	139300	51.912	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.428	75	9610	51.110	ug/l	100
93) 1,2,4-Trichlorobenzene	15.075	180	94165	56.569	ug/l	100
94) Hexachlorobutadiene	15.181	225	54487	57.902	ug/l	100
95) Naphthalene	15.310	128	182353	60.178	ug/l	100
96) 1,2,3-Trichlorobenzene	15.498	180	82005	56.350	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD090922\
Data File : VD074245.D
Acq On : 09 Sep 2022 16:55
Operator : VA/SY
Sample : VSTDICCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050


Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022

Quant Time: Sep 10 02:46:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 02:44:06 2022
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.869 min Scan# 1068
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion:168 Resp: 12712
Ion Ratio Lower Upper
168 100
99 53.8 43.0 64.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#2
Dichlorodifluoromethane
Concen: 44.557 ug/l
RT: 1.928 min Scan# 58
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 85 Resp: 46670
Ion Ratio Lower Upper
85 100
87 35.8 17.9 53.7





#3
Chloromethane
Concen: 53.209 ug/l
RT: 2.140 min Scan# 94
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 50 Resp: 40669
Ion Ratio Lower Upper
50 100
52 32.3 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#4
Vinyl Chloride
Concen: 48.149 ug/l
RT: 2.275 min Scan# 117
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Tgt Ion: 62 Resp: 37436
Ion Ratio Lower Upper
62 100
64 33.7 27.0 40.4





#5
Bromomethane
Concen: 39.412 ug/l
RT: 2.675 min Scan# 185
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 94 Resp: 1922
Ion Ratio Lower Upper
94 100
96 93.3 74.6 112.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#6
Chloroethane
Concen: 42.601 ug/l
RT: 2.822 min Scan# 210
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 64 Resp: 21014
Ion Ratio Lower Upper
64 100
66 30.7 24.6 36.8





#7
 Trichlorofluoromethane
 Concen: 45.475 ug/l
 RT: 3.158 min Scan# 20
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050



Tgt Ion:101 Resp: 102500
 Ion Ratio Lower Upper
 101 100
 103 63.3 50.6 76.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#8
 Diethyl Ether
 Concen: 54.431 ug/l
 RT: 3.587 min Scan# 340
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Tgt Ion: 74 Resp: 34120
 Ion Ratio Lower Upper
 74 100
 45 95.7 47.9 143.6









#9
1,1,2-Trichlorotrifluoroethane
Concen: 48.366 ug/l
RT: 3.952 min Scan# 402
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion:101 Resp: 66745
Ion Ratio Lower Upper
101 100
85 43.7 35.0 52.4
151 74.6 59.7 89.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#10
Methyl Iodide
Concen: 56.077 ug/l
RT: 4.152 min Scan# 436
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion:142 Resp: 84871
Ion Ratio Lower Upper
142 100
127 43.2 34.6 51.8
141 14.6 11.7 17.5





#11
Tert butyl alcohol
Concen: 282.143 ug/l
RT: 5.063 min Scan# 591
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 59 Resp: 23730
Ion Ratio Lower Upper
59 100
57 9.6 7.7 11.5

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#12
1,1-Dichloroethene
Concen: 52.010 ug/l
RT: 3.928 min Scan# 398
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Tgt Ion: 96 Resp: 64621
Ion Ratio Lower Upper
96 100
61 152.9 122.3 183.5
98 63.4 50.7 76.1





#13
Acrolein
Concen: 536.906 ug/l
RT: 3.787 min Scan# 31
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 56 Resp: 27579
Ion Ratio Lower Upper
56 100
55 71.8 57.4 86.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#14
Allyl chloride
Concen: 51.753 ug/l
RT: 4.552 min Scan# 504
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 41 Resp: 100394
Ion Ratio Lower Upper
41 100
39 66.0 52.8 79.2
76 41.1 32.9 49.3





#15
Acrylonitrile
Concen: 281.624 ug/l
RT: 5.252 min Scan# 61
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 53 Resp: 82920
Ion Ratio Lower Upper
53 100
52 81.6 65.3 97.9
51 36.3 29.0 43.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#16
Acetone
Concen: 294.006 ug/l
RT: 4.016 min Scan# 413
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 43 Resp: 73902
Ion Ratio Lower Upper
43 100
58 33.0 26.4 39.6





#17
Carbon Disulfide
Concen: 51.254 ug/l
RT: 4.258 min Scan# 454
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 76 Resp: 18228
Ion Ratio Lower Upper
76 100
78 9.0 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#18
Methyl Acetate
Concen: 50.907 ug/l
RT: 4.558 min Scan# 505
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion: 43 Resp: 40246
Ion Ratio Lower Upper
43 100
74 24.1 19.3 28.9





#19

Methyl tert-butyl Ether

Concen: 53.553 ug/l

RT: 5.316 min Scan# 61

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 73 Resp: 161772

Ion Ratio Lower Upper

73 100





57 22.2 17.8 26.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#20

Methylene Chloride

Concen: 40.719 ug/l

RT: 4.793 min Scan# 545

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion: 84 Resp: 78281




Ion Ratio Lower Upper


84 100

49 110.5 88.4 132.6

51 35.0 28.0 42.0


86 58.9 47.1 70.7





#21
trans-1,2-Dichloroethene
Concen: 50.870 ug/l
RT: 5.305 min Scan# 612
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 96 Resp: 73308
Ion Ratio Lower Upper
96 100
61 131.4 105.1 157.7
98 59.3 47.4 71.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#22
Diisopropyl ether
Concen: 54.736 ug/l
RT: 6.210 min Scan# 786
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion: 45 Resp: 221697
Ion Ratio Lower Upper
45 100
43 48.6 38.9 58.3
87 26.2 21.0 31.4
59 12.0 9.6 14.4









#23
 Vinyl Acetate
 Concen: 269.883 ug/l
 RT: 6.152 min Scan# 776
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050

Tgt Ion: 43 Resp: 550952
 Ion Ratio Lower Upper
 43 100
 86 10.8 8.6 13.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#24
 1,1-Dichloroethane
 Concen: 51.124 ug/l
 RT: 6.105 min Scan# 768
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55


Tgt Ion: 63 Resp: 135216
 Ion Ratio Lower Upper
 63 100
 98 5.6 2.8 8.4
 100 4.4 2.2 6.6





#25
2-Butanone
Concen: 279.615 ug/l
RT: 7.075 min Scan# 911
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 10911
Ion Ratio Lower Upper
43 100
72 27.2 21.8 32.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#26
2,2-Dichloropropane
Concen: 49.208 ug/l
RT: 7.069 min Scan# 932
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Tgt Ion: 77 Resp: 116501
Ion Ratio Lower Upper
77 100
97 22.9 11.5 34.4





#27
 cis-1,2-Dichloroethene
 Concen: 51.530 ug/l
 RT: 7.075 min Scan# 91
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050



Tgt Ion: 96 Resp: 84292
 Ion Ratio Lower Upper
 96 100
 61 144.7 0.0 289.4
 98 65.1 0.0 130.2




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#28
 Bromochloromethane
 Concen: 55.544 ug/l
 RT: 7.422 min Scan# 992
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Tgt Ion: 49 Resp: 45476
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 3.6
 130 77.7 62.2 93.2





#29

Tetrahydrofuran

Concen: 282.633 ug/l

RT: 7.440 min Scan# 995

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion: 42 Resp: 66459
 Ion Ratio Lower Upper
 42 100
 72 45.1 36.1 54.1
 71 42.6 34.1 51.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#30

Chloroform

Concen: 48.103 ug/l




RT: 7.587 min Scan# 1020


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion: 83 Resp: 132387
 Ion Ratio Lower Upper
 83 100
 85 70.9 56.7 85.1





#31

Cyclohexane

Concen: 51.471 ug/l

RT: 7.869 min Scan# 1068

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 56 Resp: 117478

Ion Ratio Lower Upper

56 100

69 32.1 25.7 38.5





84 92.4 73.9 110.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#32

1,1,1-Trichloroethane

Concen: 46.100 ug/l

RT: 7.787 min Scan# 1054

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55




Tgt Ion: 97 Resp: 117026


Ion Ratio Lower Upper

97 100

99 64.0 51.2 76.8

61 42.3 33.8 50.8





#33

1,2-Dichloroethane-d4

Concen: 48.941 ug/l

RT: 8.222 min Scan# 1128

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 65 Resp: 6388

Ion Ratio Lower Upper

65 100





67 55.3 0.0 110.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.769 min Scan# 1221

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55




Tgt Ion:114 Resp: 206717


Ion Ratio Lower Upper

114 100

63 18.8 0.0 37.6

88 14.2 0.0 28.4





#35

Dibromofluoromethane

Concen: 49.087 ug/l

RT: 7.799 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion:	113	Resp:	6765
Ion Ratio	Lower	Upper	
113	100		
111	103.8	83.0	124.6
192	20.2	16.2	24.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#36

1,1-Dichloropropene

Concen: 50.717 ug/l




RT: 8.004 min Scan# 1091


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion:	75	Resp:	103980
Ion Ratio	Lower	Upper	
75	100		
110	35.0	17.5	52.5
77	31.0	24.8	37.2





#37

Ethyl Acetate

Concen: 55.100 ug/l

RT: 7.163 min Scan# 948

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 43 Resp: 4579





Ion	Ratio	Lower	Upper
43	100		
61	15.1	12.1	18.1
70	10.6	8.5	12.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#38

Carbon Tetrachloride

Concen: 45.193 ug/l

RT: 7.981 min Scan# 1087




Delta R.T. 0.000 min


Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion:117 Resp: 97839

Ion	Ratio	Lower	Upper
117	100		
119	98.5	78.8	118.2
121	29.9	23.9	35.9





#39

Methylcyclohexane

Concen: 54.504 ug/l

RT: 9.263 min Scan# 1305

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion: 83	Resp: 127970
Ion Ratio	Lower Upper
83 100	
55 71.5	57.2 85.8
98 47.9	38.3 57.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#40

Benzene

Concen: 50.507 ug/l




RT: 8.240 min Scan# 1131


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55


Tgt Ion: 78	Resp: 293482
Ion Ratio	Lower Upper
78 100	
77 24.3	19.4 29.2





#41
Methacrylonitrile
Concen: 57.864 ug/l
RT: 7.399 min Scan# 988
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 41 Resp: 2931
Ion Ratio Lower Upper
41 100
39 49.7 39.8 59.6
67 70.1 56.1 84.1
52 27.5 22.0 33.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#42
1,2-Dichloroethane
Concen: 45.647 ug/l
RT: 8.316 min Scan# 1144
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Tgt Ion: 62 Resp: 77202
Ion Ratio Lower Upper
62 100
98 10.7 0.0 21.4





#43
Isopropyl Acetate
Concen: 53.816 ug/l
RT: 8.352 min Scan# 1150
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 43 Resp: 90110
Ion Ratio Lower Upper
43 100
61 30.6 24.5 36.7
87 13.5 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#44
Trichloroethene
Concen: 48.428 ug/l
RT: 9.022 min Scan# 1264
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion:130 Resp: 81326
Ion Ratio Lower Upper
130 100
95 101.1 0.0 202.2





#45

1,2-Dichloropropane

Concen: 51.800 ug/l

RT: 9.299 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 63 Resp: 75938

Ion Ratio Lower Upper

63 100





65 30.6 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#46

Dibromomethane

Concen: 49.616 ug/l

RT: 9.387 min Scan# 1326

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55




Tgt Ion: 93 Resp: 40098


Ion Ratio Lower Upper

93 100

95 84.6 67.7 101.5


174 99.5 79.6 119.4





#47
Bromodichloromethane
Concen: 48.138 ug/l
RT: 9.575 min Scan# 1358
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion: 83 Resp: 10093
Ion Ratio Lower Upper
83 100
85 66.8 53.4 80.2
127 8.6 6.9 10.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#48
Methyl methacrylate
Concen: 55.689 ug/l
RT: 9.375 min Scan# 1324
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion: 41 Resp: 42641
Ion Ratio Lower Upper
41 100
69 90.9 72.7 109.1
39 52.0 41.6 62.4





#49

1,4-Dioxane

Concen: 1118.293 ug/l

RT: 9.375 min Scan# 1324

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 88 Resp: 10548

Ion Ratio Lower Upper

88 100

43 33.7 27.0 40.4





58 67.7 54.2 81.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#50

Toluene-d8

Concen: 57.175 ug/l

RT: 10.263 min Scan# 1475

Delta R.T. 0.000 min

Lab File: VD074245.D




Acq: 09 Sep 2022 16:55


Tgt Ion: 98 Resp: 262616

Ion Ratio Lower Upper

98 100


100 64.8 51.8 77.8





#51
4-Methyl-2-Pentanone
Concen: 279.250 ug/l
RT: 10.151 min Scan# 1456
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 23331
Ion Ratio Lower Upper
43 100
58 41.6 33.3 49.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#52
Toluene
Concen: 49.873 ug/l
RT: 10.328 min Scan# 1486
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 92 Resp: 189516
Ion Ratio Lower Upper
92 100
91 174.1 139.3 208.9





#53
 t-1,3-Dichloropropene
 Concen: 51.012 ug/l
 RT: 10.546 min Scan# 1523
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55




Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050



Tgt Ion: 75 Resp: 9752
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.7 38.5




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#54
 cis-1,3-Dichloropropene
 Concen: 51.409 ug/l
 RT: 10.010 min Scan# 1432
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55


Tgt Ion: 75 Resp: 116661
 Ion Ratio Lower Upper
 75 100
 77 31.7 25.4 38.0
 39 39.5 31.6 47.4





#55
1,1,2-Trichloroethane
Concen: 50.529 ug/l
RT: 10.722 min Scan# 1553
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion:	97	Resp:	56474
Ion	Ratio	Lower	Upper
97	100		
83	85.0	68.0	102.0
85	57.3	45.8	68.8
99	63.4	50.7	76.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#56
Ethyl methacrylate
Concen: 59.253 ug/l
RT: 10.587 min Scan# 1530
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion:	69	Resp:	76791
Ion	Ratio	Lower	Upper
69	100		
41	61.8	49.4	74.2
39	30.3	24.2	36.4





#57

1,3-Dichloropropane

Concen: 52.433 ug/l

RT: 10.869 min Scan# 1578

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050

Tgt Ion: 76 Resp: 9801

Ion Ratio Lower Upper

76 100





78 31.8 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#58

2-Chloroethyl Vinyl ether

Concen: 163.069 ug/l

RT: 9.863 min Scan# 1407

Delta R.T. 0.000 min

Lab File: VD074245.D




Acq: 09 Sep 2022 16:55


Tgt Ion: 63 Resp: 60136

Ion Ratio Lower Upper

63 100


106 27.4 21.9 32.9





#59
2-Hexanone
Concen: 289.845 ug/l
RT: 10.910 min Scan# 1585
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion: 43 Resp: 16321
Ion Ratio Lower Upper
43 100
58 57.2 28.6 85.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#60
Dibromochloromethane
Concen: 47.601 ug/l
RT: 11.063 min Scan# 1611
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:129 Resp: 67886
Ion Ratio Lower Upper
129 100
127 77.9 39.0 116.9









#61
1,2-Dibromoethane
Concen: 49.466 ug/l
RT: 11.169 min Scan# 1629
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion:107 Resp: 53308
Ion Ratio Lower Upper
107 100
109 92.1 73.7 110.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#62
4-Bromofluorobenzene
Concen: 48.226 ug/l
RT: 12.557 min Scan# 1865
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 95 Resp: 83434
Ion Ratio Lower Upper
95 100
174 89.6 0.0 179.2
176 85.7 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.569 min Scan# 1697
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion:117 Resp: 186309
Ion Ratio Lower Upper
117 100
82 55.2 44.2 66.2
119 31.7 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#64
Tetrachloroethene
Concen: 49.167 ug/l
RT: 10.798 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Tgt Ion:164 Resp: 65942
Ion Ratio Lower Upper
164 100
166 117.3 93.8 140.8
129 85.4 68.3 102.5
131 83.9 67.1 100.7





#65
Chlorobenzene
Concen: 49.897 ug/l
RT: 11.598 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion:112 Resp: 20103
Ion Ratio Lower Upper
112 100
114 31.9 25.5 38.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#66
1,1,1,2-Tetrachloroethane
Concen: 47.988 ug/l
RT: 11.669 min Scan# 1714
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:131 Resp: 72307
Ion Ratio Lower Upper
131 100
133 95.6 47.8 143.4
119 66.3 33.1 99.5









#67
Ethyl Benzene
Concen: 53.827 ug/l
RT: 11.669 min Scan# 1714
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 37314
Ion Ratio Lower Upper
91 100
106 31.0 24.8 37.2




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#68
m/p-Xylenes
Concen: 104.023 ug/l
RT: 11.781 min Scan# 1733
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:106 Resp: 283947
Ion Ratio Lower Upper
106 100
91 198.2 158.6 237.8









#69
o-Xylene
Concen: 53.307 ug/l
RT: 12.110 min Scan# 1789
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion:106 Resp: 135088
Ion Ratio Lower Upper
106 100
91 208.0 104.0 312.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#70
Styrene
Concen: 52.602 ug/l
RT: 12.122 min Scan# 1791
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:104 Resp: 227820
Ion Ratio Lower Upper
104 100
78 48.0 38.4 57.6
103 55.1 44.1 66.1









#71
Bromoform
Concen: 50.020 ug/l
RT: 12.287 min Scan# 1819
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion:173 Resp: 4136
Ion Ratio Lower Upper
173 100
175 48.4 24.2 72.6
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.498 min Scan# 2025
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion:152 Resp: 88587
Ion Ratio Lower Upper
152 100
115 56.8 28.4 85.2
150 175.5 0.0 351.0





#73
Isopropylbenzene
Concen: 55.774 ug/l
RT: 12.404 min Scan# 1839
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion:105 Resp: 35918
Ion Ratio Lower Upper
105 100
120 26.4 13.2 39.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#74
N-ethyl acetate
Concen: 60.708 ug/l
RT: 12.216 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion: 43 Resp: 87417
Ion Ratio Lower Upper
43 100
70 43.2 34.6 51.8
55 22.7 18.2 27.2
61 25.8 20.6 31.0









#75
 1,1,2,2-Tetrachloroethane
 Concen: 56.676 ug/l
 RT: 12.657 min Scan# 1882
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050

Tgt Ion: 83 Resp: 66485
 Ion Ratio Lower Upper
 83 100
 131 9.8 4.9 14.7
 85 64.6 32.3 96.9




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#76
 1,2,3-Trichloropropane
 Concen: 54.106 ug/l m
 RT: 12.710 min Scan# 1891
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Tgt Ion: 75 Resp: 48155
 Ion Ratio Lower Upper
 75 100
 77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 51.304 ug/l

RT: 12.687 min Scan# 1887

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion:	156	Resp:	78038
Ion Ratio	Lower	Upper	
156	100		
77	182.2	91.1	273.3
158	98.8	49.4	148.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#78

n-propylbenzene

Concen: 55.412 ug/l




RT: 12.745 min Scan# 1897


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion:	91	Resp:	436542
Ion Ratio	Lower	Upper	
91	100		
120	23.1	11.6	34.7









#79
2-Chlorotoluene
Concen: 52.703 ug/l
RT: 12.834 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 23822
Ion Ratio Lower Upper
91 100
126 34.8 17.3 51.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#80
1,3,5-Trimethylbenzene
Concen: 53.293 ug/l
RT: 12.887 min Scan# 1921
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:105 Resp: 292246
Ion Ratio Lower Upper
105 100
120 49.3 24.6 74.0









#81
trans-1,4-Dichloro-2-butene
Concen: 53.507 ug/l
RT: 12.451 min Scan# 1847
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion: 75 Resp: 1808
Ion Ratio Lower Upper
75 100
53 87.8 0.0 0.0
89 42.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#82
4-Chlorotoluene
Concen: 50.656 ug/l
RT: 12.928 min Scan# 1928
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55


Tgt Ion: 91 Resp: 240486
Ion Ratio Lower Upper
91 100
126 34.4 17.2 51.6





#83
tert-Butylbenzene
Concen: 55.223 ug/l
RT: 13.151 min Scan# 1966
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050




Tgt Ion:119 Resp: 25811
Ion Ratio Lower Upper
119 100
91 62.6 31.3 93.9
134 26.7 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022




#84
1,2,4-Trimethylbenzene
Concen: 53.887 ug/l
RT: 13.192 min Scan# 1973
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55



Tgt Ion:105 Resp: 288287
Ion Ratio Lower Upper
105 100
120 46.2 23.1 69.3









#85
 sec-Butylbenzene
 Concen: 55.364 ug/l
 RT: 13.328 min Scan# 1996
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Instrument :
 MSVOA_D
 ClientSampleId :
 VSTDICCC050

Tgt Ion:105 Resp: 390249
 Ion Ratio Lower Upper
 105 100
 134 19.9 10.0 29.8




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 09/12/2022
 Supervised By :Mahesh Dadoda 09/12/2022



#86
 p-Isopropyltoluene
 Concen: 55.670 ug/l
 RT: 13.440 min Scan# 2015
 Delta R.T. 0.000 min
 Lab File: VD074245.D
 Acq: 09 Sep 2022 16:55

Tgt Ion:119 Resp: 322343
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.1 39.3
 91 22.3 11.2 33.5





#87

1,3-Dichlorobenzene

Concen: 51.620 ug/l

RT: 13.440 min Scan# 2015

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion:146	Resp: 15844
Ion Ratio	Lower Upper
146	100
111	39.2 18.1 54.4
148	63.4 31.7 95.1

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#88

1,4-Dichlorobenzene

Concen: 50.531 ug/l




RT: 13.522 min Scan# 2029


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Tgt Ion:146	Resp: 156276
Ion Ratio	Lower Upper
146	100
111	36.8 18.4 55.2
148	64.3 32.1 96.5









#89
n-Butylbenzene
Concen: 56.623 ug/l
RT: 13.769 min Scan# 2071
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion: 91 Resp: 31363
Ion Ratio Lower Upper
91 100
92 53.1 26.6 79.7
134 26.1 13.1 39.1




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#90
Hexachloroethane
Concen: 50.262 ug/l
RT: 14.034 min Scan# 2116
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:117 Resp: 58906
Ion Ratio Lower Upper
117 100
201 75.9 38.0 113.9









#91
1,2-Dichlorobenzene
Concen: 51.912 ug/l
RT: 13.810 min Scan# 2078
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050

Tgt Ion:146 Resp: 139300
Ion Ratio Lower Upper
146 100
111 38.9 19.4 58.4
148 63.0 31.5 94.5




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 51.110 ug/l
RT: 14.428 min Scan# 2183
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion: 75 Resp: 9610
Ion Ratio Lower Upper
75 100
155 98.4 49.2 147.6
157 124.1 62.1 186.2





#93

1,2,4-Trichlorobenzene

Concen: 56.569 ug/l

RT: 15.075 min Scan# 2293

Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55

Instrument :

MSVOA_D

ClientSampleId :

VSTDICCC050





Tgt Ion:	180	Resp:	9416
Ion Ratio	Lower	Upper	
180	100		
182	95.3	47.6	142.9
145	28.9	14.4	43.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 09/12/2022

Supervised By :Mahesh Dadoda 09/12/2022



#94

Hexachlorobutadiene

Concen: 57.902 ug/l




RT: 15.181 min Scan# 2311


Delta R.T. 0.000 min

Lab File: VD074245.D

Acq: 09 Sep 2022 16:55


Tgt Ion:	225	Resp:	54487
Ion Ratio	Lower	Upper	
225	100		
223	64.7	32.4	97.0
227	65.4	32.7	98.1





#95
Naphthalene
Concen: 60.178 ug/l
RT: 15.310 min Scan# 2333
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55




Instrument :
MSVOA_D
ClientSampleId :
VSTDICCC050



Tgt Ion:128 Resp: 18235
Ion Ratio Lower Upper
128 100
127 13.6 10.9 16.3
129 11.0 8.8 13.2

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 09/12/2022
Supervised By :Mahesh Dadoda 09/12/2022



#96
1,2,3-Trichlorobenzene
Concen: 56.350 ug/l
RT: 15.498 min Scan# 2365
Delta R.T. 0.000 min
Lab File: VD074245.D
Acq: 09 Sep 2022 16:55

Tgt Ion:180 Resp: 82005
Ion Ratio Lower Upper
180 100
182 95.0 47.5 142.5
145 31.5 15.8 47.3

