


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD091020\
 Data File : VD066627.D
 Acq On : 10 Sep 2020 21:11
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Title : SW846 8260
 Last Update : Fri Sep 04 01:57:04 2020



AutoFind: Scans 1821, 1822, 1823; Background Corrected with Scan 1811

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.0	37093	PASS
75	95	30	60	54.6	87946	PASS
95	95	100	100	100.0	161133	PASS
96	95	5	9	6.5	10516	PASS
173	174	0.00	2	0.5	616	PASS
174	95	50	100	78.3	126210	PASS
175	174	5	9	8.3	10424	PASS
176	174	95	101	95.5	120544	PASS
177	176	5	9	6.4	7659	PASS