

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070298.D
 Acq On : 10 Sep 2021 14:25
 Operator : VA/SY
 Sample : M3691-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-03-090921

Quant Time: Sep 11 02:25:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	22686	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	44031	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	46328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	22272	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	24735	84.914	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 169.820%#		
35) Dibromofluoromethane	7.909	113	17340	55.325	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 110.660%		
50) Toluene-d8	10.332	98	56088	49.289	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 98.580%		
62) 4-Bromofluorobenzene	12.626	95	19129	48.099	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 96.200%		
Target Compounds						
						Qvalue
16) Acetone	4.156	43	1937	37.246	ug/l #	83
20) Methylene Chloride	4.973	84	4281	6.956	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070298.D
Acq On : 10 Sep 2021 14:25
Operator : VA/SY
Sample : M3691-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-03-090921

Quant Time: Sep 11 02:25:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

