

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070302.D
 Acq On : 10 Sep 2021 16:19
 Operator : VA/SY
 Sample : M3684-03
 Misc : 5.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUP-210908A

Quant Time: Sep 11 02:26:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	279777	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	653844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	698988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	305111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	238778	66.468	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 132.940%	
35) Dibromofluoromethane	7.914	113	249623	53.635	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.260%	
50) Toluene-d8	10.338	98	876432	51.866	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 103.740%	
62) 4-Bromofluorobenzene	12.626	95	315208	53.374	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 106.740%	
Target Compounds						
32) 1,1,1-Trichloroethane	7.897	97	18748	2.997	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070302.D
Acq On : 10 Sep 2021 16:19
Operator : VA/SY
Sample : M3684-03
Misc : 5.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUP-210908A

Quant Time: Sep 11 02:26:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

