

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070304.D
 Acq On : 10 Sep 2021 17:16
 Operator : VA/SY
 Sample : M3684-05
 Misc : 10.90G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DUP-210908B

Quant Time: Sep 11 02:27:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	4238	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.861	114	8999	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	10012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	3117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	4706	86.481	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 172.960%#	
35) Dibromofluoromethane	7.908	113	3725	58.152	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.300%	
50) Toluene-d8	10.338	98	10790	46.394	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 92.780%	
62) 4-Bromofluorobenzene	12.620	95	3648	44.881	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.760%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070304.D
Acq On : 10 Sep 2021 17:16
Operator : VA/SY
Sample : M3684-05
Misc : 10.90G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DUP-210908B

Quant Time: Sep 11 02:27:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

