

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070307.D
 Acq On : 10 Sep 2021 18:42
 Operator : VA/SY
 Sample : M3684-08
 Misc : 8.95G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A019145

Quant Time: Sep 11 02:28:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	254435	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	618016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	692069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	295632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	230525	70.562	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 141.120%	
35) Dibromofluoromethane	7.914	113	243974	55.460	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 110.920%	
50) Toluene-d8	10.338	98	854762	53.516	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.040%	
62) 4-Bromofluorobenzene	12.626	95	304178	54.492	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.980%	
Target Compounds						
						Qvalue
16) Acetone	4.156	43	15803	27.094	ug/l	97
32) 1,1,1-Trichloroethane	7.903	97	14267	2.508	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070307.D
Acq On : 10 Sep 2021 18:42
Operator : VA/SY
Sample : M3684-08
Misc : 8.95G/5.00ml/MSVOA_D/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A019145

Quant Time: Sep 11 02:28:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

