

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070309.D
 Acq On : 10 Sep 2021 19:39
 Operator : VA/SY
 Sample : M3684-10
 Misc : 5.88G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A017015

Quant Time: Sep 11 02:28:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	242114	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	597614	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	652272	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	251969	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	222036	71.422	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	142.840%
35) Dibromofluoromethane	7.914	113	234258	55.069	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.140%
50) Toluene-d8	10.338	98	830152	53.749	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.500%
62) 4-Bromofluorobenzene	12.626	95	277518	51.413	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070309.D
Acq On : 10 Sep 2021 19:39
Operator : VA/SY
Sample : M3684-10
Misc : 5.88G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A017015

Quant Time: Sep 11 02:28:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

