

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070311.D
 Acq On : 10 Sep 2021 20:36
 Operator : VA/SY
 Sample : M3684-12
 Misc : 5.91G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 F005130

Quant Time: Sep 11 02:29:26 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	248389	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	606801	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	720374	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	300894	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	219404	68.792	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	137.580%
35) Dibromofluoromethane	7.914	113	232866	53.913	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.820%
50) Toluene-d8	10.338	98	852785	54.379	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	108.760%
62) 4-Bromofluorobenzene	12.626	95	311044	56.752	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	113.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070311.D
Acq On : 10 Sep 2021 20:36
Operator : VA/SY
Sample : M3684-12
Misc : 5.91G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
F005130

Quant Time: Sep 11 02:29:26 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

