

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070313.D
 Acq On : 10 Sep 2021 21:34
 Operator : VA/SY
 Sample : M3684-14
 Misc : 6.54G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A008015

Quant Time: Sep 11 02:30:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	253087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	609703	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	671163	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	272904	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	216881	66.739	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	133.480%
35) Dibromofluoromethane	7.914	113	232665	53.610	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.220%
50) Toluene-d8	10.338	98	845440	53.654	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.300%
62) 4-Bromofluorobenzene	12.626	95	288784	52.440	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	104.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070313.D
Acq On : 10 Sep 2021 21:34
Operator : VA/SY
Sample : M3684-14
Misc : 6.54G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A008015

Quant Time: Sep 11 02:30:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

