

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070294.D
 Acq On : 10 Sep 2021 12:24
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 11 02:23:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	90	0.00
2 T	Dichlorodifluoromethane	0.574	0.477	16.9	91	0.00
3 P	Chloromethane	0.771	0.628	18.5	89	0.00
4 C	Vinyl Chloride	0.787	0.686	12.8#	90	0.00
5 T	Bromomethane	0.580	0.498	14.1	100	0.00
6 T	Chloroethane	0.541	0.476	12.0	97	0.00
7 T	Trichlorofluoromethane	1.041	0.933	10.4	87	0.00
8 T	Diethyl Ether	0.288	0.239	17.0	75	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.646	0.591	8.5	88	0.00
10 T	Methyl Iodide	0.564	0.505	10.5	72	0.00
11 T	Tert butyl alcohol	0.073	0.040	45.2#	72	-0.01
12 CM	1,1-Dichloroethene	0.582	0.512	12.0#	82	0.00
13 T	Acrolein	0.032	0.037	-15.6	111	0.00
14 T	Allyl chloride	0.855	0.754	11.8	84	0.00
15 T	Acrylonitrile	0.137	0.116	15.3	81	0.00
16 T	Acetone	0.115	0.111	3.5	97	0.00
17 T	Carbon Disulfide	2.148	1.904	11.4	86	0.00
18 T	Methyl Acetate	0.344	0.274	20.3	84	0.00
19 T	Methyl tert-butyl Ether	1.258	1.103	12.3	77	0.00
20 T	Methylene Chloride	0.913	0.706	22.7	88	0.00
21 T	trans-1,2-Dichloroethene	0.693	0.630	9.1	84	0.00
22 T	Diisopropyl ether	1.831	1.729	5.6	87	0.00
23 T	Vinyl Acetate	0.848	0.793	6.5	86	0.00
24 P	1,1-Dichloroethane	1.295	1.147	11.4	87	0.00
25 T	2-Butanone	0.153	0.136	11.1	88	0.00
26 T	2,2-Dichloropropane	1.047	0.947	9.6	89	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.665	10.6	82	0.00
28 T	Bromochloromethane	0.514	0.527	-2.5	100	0.00
29 T	Tetrahydrofuran	0.098	0.084	14.3	83	0.00
30 C	Chloroform	1.328	1.169	12.0#	86	0.00
31 T	Cyclohexane	1.106	1.004	9.2	88	0.00
32 T	1,1,1-Trichloroethane	1.118	0.995	11.0	86	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.583	9.2	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	89	0.00
35 S	Dibromofluoromethane	0.356	0.347	2.5	90	0.00
36 T	1,1-Dichloropropene	0.534	0.523	2.1	88	0.00
37 T	Ethyl Acetate	0.200	0.180	10.0	87	0.00
38 T	Carbon Tetrachloride	0.529	0.499	5.7	85	0.00
39 T	Methylcyclohexane	0.597	0.597	0.0	84	0.00
40 TM	Benzene	1.566	1.484	5.2	84	0.00
41 T	Methacrylonitrile	0.105	0.092	12.4	85	0.00
42 TM	1,2-Dichloroethane	0.419	0.387	7.6	87	0.00
43 T	Isopropyl Acetate	0.370	0.331	10.5	85	0.00
44 TM	Trichloroethene	0.374	0.348	7.0	82	0.00
45 C	1,2-Dichloropropane	0.411	0.385	6.3#	87	0.00
46 T	Dibromomethane	0.212	0.192	9.4	84	0.00
47 T	Bromodichloromethane	0.540	0.499	7.6	85	0.00
48 T	Methyl methacrylate	0.171	0.164	4.1	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
 Data File : VD070294.D
 Acq On : 10 Sep 2021 12:24
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 11 02:23:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	76	0.00
50 S	Toluene-d8	1.292	1.330	-2.9	90	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.178	6.8	85	0.00
52 CM	Toluene	0.946	0.930	1.7#	83	0.00
53 T	t-1,3-Dichloropropene	0.464	0.435	6.3	83	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.532	5.8	86	0.00
55 T	1,1,2-Trichloroethane	0.291	0.264	9.3	81	0.00
56 T	Ethyl methacrylate	0.318	0.293	7.9	76	0.00
57 T	1,3-Dichloropropane	0.496	0.457	7.9	83	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.175	-14.4	101	0.00
59 T	2-Hexanone	0.128	0.124	3.1	86	0.00
60 T	Dibromochloromethane	0.339	0.308	9.1	79	0.00
61 T	1,2-Dibromoethane	0.268	0.241	10.1	81	0.00
62 S	4-Bromofluorobenzene	0.452	0.448	0.9	88	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.304	0.295	3.0	83	0.00
65 PM	Chlorobenzene	1.036	0.972	6.2	82	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.354	6.6	83	0.00
67 C	Ethyl Benzene	1.786	1.784	0.1#	83	0.00
68 T	m/p-Xylenes	0.690	0.699	-1.3	82	0.00
69 T	o-Xylene	0.618	0.624	-1.0	81	0.00
70 T	Styrene	1.080	1.113	-3.1	81	0.00
71 P	Bromoform	0.187	0.168	10.2	78	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	3.506	3.399	3.1	82	0.00
74 T	N-amyl acetate	0.764	0.684	10.5	85	0.00
75 P	1,1,2,2-Tetrachloroethane	0.751	0.640	14.8	82	0.00
76 T	1,2,3-Trichloropropane	0.506	0.441	12.8	82	0.00
77 T	Bromobenzene	0.809	0.733	9.4	81	0.00
78 T	n-propylbenzene	4.551	4.533	0.4	86	0.00
79 T	2-Chlorotoluene	2.653	2.542	4.2	84	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.005	0.7	83	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.192	5.0	88	0.00
82 T	4-Chlorotoluene	2.799	2.687	4.0	84	0.00
83 T	tert-Butylbenzene	2.427	2.369	2.4	82	0.00
84 T	1,2,4-Trimethylbenzene	2.980	3.009	-1.0	84	0.00
85 T	sec-Butylbenzene	3.858	3.861	-0.1	85	0.00
86 T	p-Isopropyltoluene	3.100	3.141	-1.3	84	0.00
87 T	1,3-Dichlorobenzene	1.684	1.546	8.2	82	0.00
88 T	1,4-Dichlorobenzene	1.691	1.529	9.6	83	0.00
89 T	n-Butylbenzene	3.091	3.096	-0.2	86	0.00
90 T	Hexachloroethane	0.710	0.652	8.2	88	0.00
91 T	1,2-Dichlorobenzene	1.486	1.341	9.8	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.093	17.0	81	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.721	10.1	80	0.00
94 T	Hexachlorobutadiene	0.492	0.454	7.7	89	0.00
95 T	Naphthalene	1.458	1.258	13.7	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091021\
Data File : VD070294.D
Acq On : 10 Sep 2021 12:24
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 11 02:23:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.708	0.625	11.7	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6