

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091120\
 Data File : VD066654.D
 Acq On : 11 Sep 2020 17:52
 Operator : VA/SY
 Sample : L3981-05
 Misc : 10.71G/5.00ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B18-4-6

Manual Integrations
 APPROVED

MMDadoda
 9/14/2020 12:07:01 PM

Quant Time: Sep 21 00:56:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 04 01:57:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	371117	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	654381	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	601586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	225096	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	238490	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
35) Dibromofluoromethane	7.91	113	222114	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
50) Toluene-d8	10.34	98	794315	50.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.68%	
62) 4-Bromofluorobenzene	12.63	95	239071	45.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.40%	
Target Compounds						
21) trans-1,2-Dichloroethene	5.47	96	10748m	2.421	ug/l	
27) cis-1,2-Dichloroethene	7.21	96	386976	81.046	ug/l	89
30) Chloroform	7.71	83	11431	1.364	ug/l	94
44) Trichloroethene	9.11	130	888633	181.634	ug/l	96
64) Tetrachloroethene	10.87	164	22811581	5631.793	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091120\
Data File : VD066654.D
Acq On : 11 Sep 2020 17:52
Operator : VA/SY
Sample : L3981-05
Misc : 10.71G/5.00ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B18-4-6

Manual Integrations
APPROVED

MMDadoda
9/14/2020 12:07:01 PM

Quant Time: Sep 21 00:56:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 04 01:57:04 2020
Response via : Initial Calibration

