

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD091120\
 Data File : VD066661.D
 Acq On : 11 Sep 2020 21:04
 Operator : VA/SY
 Sample : L3981-12
 Misc : 12.16G/5.00ml/MSVOA D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 B20-7-9

Quant Time: Sep 12 02:13:14 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
 Quant Title : SW846 8260
 QLast Update : Fri Sep 04 01:57:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	360417	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.86	114	663285	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.64	117	623023	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	265341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	237900	54.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.12%	
35) Dibromofluoromethane	7.91	113	228440	51.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.10%	
50) Toluene-d8	10.34	98	826414	51.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	12.63	95	272065	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
Target Compounds						
44) Trichloroethene	9.11	130	5304	1.070	ug/l	100
64) Tetrachloroethene	10.87	164	476499	113.592	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD091120\
Data File : VD066661.D
Acq On : 11 Sep 2020 21:04
Operator : VA/SY
Sample : L3981-12
Misc : 12.16G/5.00ml/MSVOA D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
B20-7-9

Quant Time: Sep 12 02:13:14 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D090320S.M
Quant Title : SW846 8260
QLast Update : Fri Sep 04 01:57:04 2020
Response via : Initial Calibration

