

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070319.D
 Acq On : 13 Sep 2021 11:39
 Operator : VA/SY
 Sample : VD0913SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD0913SBL01

Quant Time: Sep 14 00:50:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	288923	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	672125	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	748798	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	326100	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	232867	62.770	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.540%
35) Dibromofluoromethane	7.914	113	250865	52.435	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.880%
50) Toluene-d8	10.338	98	918537	52.879	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.760%
62) 4-Bromofluorobenzene	12.626	95	342913	56.486	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	112.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070319.D
Acq On : 13 Sep 2021 11:39
Operator : VA/SY
Sample : VD0913SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD0913SBL01

Quant Time: Sep 14 00:50:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

