

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070323.D
 Acq On : 13 Sep 2021 13:35
 Operator : VA/SY
 Sample : M3685-01RE
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 327RE

Quant Time: Sep 14 00:52:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	70506	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	151825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	174755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	67886	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	79060	87.329	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 174.660%#	
35) Dibromofluoromethane	7.914	113	69983	64.757	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 129.520%	
50) Toluene-d8	10.338	98	200616	51.128	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.260%	
62) 4-Bromofluorobenzene	12.626	95	69380	50.594	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 101.180%	
Target Compounds						
16) Acetone	4.167	43	3487	21.574	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070323.D
Acq On : 13 Sep 2021 13:35
Operator : VA/SY
Sample : M3685-01RE
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
327RE

Quant Time: Sep 14 00:52:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

