

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070324.D
 Acq On : 13 Sep 2021 14:04
 Operator : VA/SY
 Sample : M3719-01
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-01-091021

Manual Integrations
 APPROVED

MMDadoda
 9/14/2021 10:45:27 AM

Quant Time: Sep 14 00:53:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	1973	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.855	114	3777	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.644	117	3873	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.555	152	1295	50.000	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	1444m	56.999	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 114.000%	
35) Dibromofluoromethane	7.908	113	1657	61.632	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 123.260%	
50) Toluene-d8	10.332	98	4918	50.382	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.760%	
62) 4-Bromofluorobenzene	12.632	95	1850	54.229	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070324.D
Acq On : 13 Sep 2021 14:04
Operator : VA/SY
Sample : M3719-01
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-01-091021

Manual Integrations
APPROVED

MMDadoda
9/14/2021 10:45:27 AM

Quant Time: Sep 14 00:53:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

