

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070336.D
 Acq On : 13 Sep 2021 19:48
 Operator : VA/SY
 Sample : M3687-25
 Misc : 5.71G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A015015

Quant Time: Sep 14 00:56:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	378424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	779224	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	902907	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	417392	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	258914	53.285	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.560%
35) Dibromofluoromethane	7.914	113	276223	49.800	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.600%
50) Toluene-d8	10.338	98	1042800	51.782	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.560%
62) 4-Bromofluorobenzene	12.626	95	435101	61.821	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	123.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070336.D
Acq On : 13 Sep 2021 19:48
Operator : VA/SY
Sample : M3687-25
Misc : 5.71G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A015015

Quant Time: Sep 14 00:56:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

