

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070338.D
 Acq On : 13 Sep 2021 20:45
 Operator : VA/SY
 Sample : M3687-27
 Misc : 4.99G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A011145

Quant Time: Sep 14 00:57:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	350654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	698924	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	800829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	279986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	229733	51.024	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.040%	
35) Dibromofluoromethane	7.914	113	249835	50.218	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.440%	
50) Toluene-d8	10.338	98	910891	50.428	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 100.860%	
62) 4-Bromofluorobenzene	12.626	95	341833	54.149	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.300%	
Target Compounds						
						Qvalue
30) Chloroform	7.714	83	16204	1.740	ug/l	90
32) 1,1,1-Trichloroethane	7.903	97	777905	99.230	ug/l	98
64) Tetrachloroethene	10.873	164	473415	97.095	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070338.D
Acq On : 13 Sep 2021 20:45
Operator : VA/SY
Sample : M3687-27
Misc : 4.99G/5.00ml/MSVOA_D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A011145

Quant Time: Sep 14 00:57:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

