

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070340.D
 Acq On : 13 Sep 2021 21:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 14 00:57:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.574	0.483	15.9	88	0.00
3 P	Chloromethane	0.771	0.752	2.5	102	0.00
4 C	Vinyl Chloride	0.787	0.745	5.3#	94	0.00
5 T	Bromomethane	0.580	0.532	8.3	102	0.00
6 T	Chloroethane	0.541	0.519	4.1	101	0.00
7 T	Trichlorofluoromethane	1.041	0.940	9.7	84	0.00
8 T	Diethyl Ether	0.288	0.304	-5.6	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.646	0.586	9.3	84	0.00
10 T	Methyl Iodide	0.564	0.603	-6.9	82	0.00
11 T	Tert butyl alcohol	0.073	0.045	38.4#	78	0.00
12 CM	1,1-Dichloroethene	0.582	0.564	3.1#	87	0.00
13 T	Acrolein	0.032	0.036	-12.5	103	0.00
14 T	Allyl chloride	0.855	0.866	-1.3	93	0.00
15 T	Acrylonitrile	0.137	0.144	-5.1	97	0.00
16 T	Acetone	0.115	0.100	13.0	84	0.00
17 T	Carbon Disulfide	2.148	2.026	5.7	88	0.00
18 T	Methyl Acetate	0.344	0.351	-2.0	103	0.00
19 T	Methyl tert-butyl Ether	1.258	1.355	-7.7	91	0.00
20 T	Methylene Chloride	0.913	0.990	-8.4	118	0.00
21 T	trans-1,2-Dichloroethene	0.693	0.708	-2.2	91	0.00
22 T	Diisopropyl ether	1.831	2.054	-12.2	99	0.00
23 T	Vinyl Acetate	0.848	0.915	-7.9	95	0.00
24 P	1,1-Dichloroethane	1.295	1.307	-0.9	95	0.00
25 T	2-Butanone	0.153	0.154	-0.7	95	0.00
26 T	2,2-Dichloropropane	1.047	0.927	11.5	84	0.00
27 T	cis-1,2-Dichloroethene	0.744	0.785	-5.5	93	0.00
28 T	Bromochloromethane	0.514	0.537	-4.5	98	0.00
29 T	Tetrahydrofuran	0.098	0.104	-6.1	98	0.00
30 C	Chloroform	1.328	1.344	-1.2#	95	0.00
31 T	Cyclohexane	1.106	1.013	8.4	85	0.00
32 T	1,1,1-Trichloroethane	1.118	1.087	2.8	90	0.00
33 S	1,2-Dichloroethane-d4	0.642	0.666	-3.7	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.356	0.365	-2.5	97	0.00
36 T	1,1-Dichloropropene	0.534	0.522	2.2	90	0.00
37 T	Ethyl Acetate	0.200	0.204	-2.0	101	0.00
38 T	Carbon Tetrachloride	0.529	0.488	7.8	86	0.00
39 T	Methylcyclohexane	0.597	0.560	6.2	81	0.00
40 TM	Benzene	1.566	1.599	-2.1	94	0.00
41 T	Methacrylonitrile	0.105	0.102	2.9	96	0.00
42 TM	1,2-Dichloroethane	0.419	0.421	-0.5	97	0.00
43 T	Isopropyl Acetate	0.370	0.364	1.6	96	0.00
44 TM	Trichloroethene	0.374	0.374	0.0	90	0.00
45 C	1,2-Dichloropropane	0.411	0.424	-3.2#	98	0.00
46 T	Dibromomethane	0.212	0.215	-1.4	97	0.00
47 T	Bromodichloromethane	0.540	0.548	-1.5	96	0.00
48 T	Methyl methacrylate	0.171	0.167	2.3	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
 Data File : VD070340.D
 Acq On : 13 Sep 2021 21:41
 Operator : VA/SY
 Sample : VSTDCCC050
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Sep 14 00:57:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	93	0.00
50 S	Toluene-d8	1.292	1.395	-8.0	97	0.00
51 T	4-Methyl-2-Pentanone	0.191	0.202	-5.8	99	0.00
52 CM	Toluene	0.946	0.980	-3.6#	90	0.00
53 T	t-1,3-Dichloropropene	0.464	0.459	1.1	90	0.00
54 T	cis-1,3-Dichloropropene	0.565	0.554	1.9	92	0.00
55 T	1,1,2-Trichloroethane	0.291	0.292	-0.3	92	0.00
56 T	Ethyl methacrylate	0.318	0.344	-8.2	92	0.00
57 T	1,3-Dichloropropane	0.496	0.513	-3.4	96	0.00
58 T	2-Chloroethyl Vinyl ether	0.153	0.162	-5.9	96	0.00
59 T	2-Hexanone	0.128	0.134	-4.7	97	0.00
60 T	Dibromochloromethane	0.339	0.346	-2.1	92	0.00
61 T	1,2-Dibromoethane	0.268	0.275	-2.6	95	0.00
62 S	4-Bromofluorobenzene	0.452	0.476	-5.3	96	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
64 T	Tetrachloroethene	0.304	0.303	0.3	91	0.00
65 PM	Chlorobenzene	1.036	1.028	0.8	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.380	-0.3	94	0.00
67 C	Ethyl Benzene	1.786	1.857	-4.0#	93	0.00
68 T	m/p-Xylenes	0.690	0.720	-4.3	90	0.00
69 T	o-Xylene	0.618	0.650	-5.2	90	0.00
70 T	Styrene	1.080	1.189	-10.1	92	0.00
71 P	Bromoform	0.187	0.187	0.0	92	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.506	3.624	-3.4	91	0.00
74 T	N-amyl acetate	0.764	0.760	0.5	98	0.00
75 P	1,1,2,2-Tetrachloroethane	0.751	0.723	3.7	97	0.00
76 T	1,2,3-Trichloropropane	0.506	0.480	5.1	93	0.00
77 T	Bromobenzene	0.809	0.809	0.0	94	0.00
78 T	n-propylbenzene	4.551	4.642	-2.0	92	0.00
79 T	2-Chlorotoluene	2.653	2.675	-0.8	92	0.00
80 T	1,3,5-Trimethylbenzene	3.026	3.177	-5.0	91	0.00
81 T	trans-1,4-Dichloro-2-butene	0.202	0.178	11.9	85	0.00
82 T	4-Chlorotoluene	2.799	2.807	-0.3	91	0.00
83 T	tert-Butylbenzene	2.427	2.520	-3.8	91	0.00
84 T	1,2,4-Trimethylbenzene	2.980	3.187	-6.9	93	0.00
85 T	sec-Butylbenzene	3.858	3.906	-1.2	90	0.00
86 T	p-Isopropyltoluene	3.100	3.237	-4.4	90	0.00
87 T	1,3-Dichlorobenzene	1.684	1.678	0.4	92	0.00
88 T	1,4-Dichlorobenzene	1.691	1.646	2.7	93	0.00
89 T	n-Butylbenzene	3.091	3.042	1.6	88	0.00
90 T	Hexachloroethane	0.710	0.673	5.2	95	0.00
91 T	1,2-Dichlorobenzene	1.486	1.473	0.9	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.112	0.106	5.4	97	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.804	-0.2	94	0.00
94 T	Hexachlorobutadiene	0.492	0.452	8.1	93	0.00
95 T	Naphthalene	1.458	1.510	-3.6	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091321\
Data File : VD070340.D
Acq On : 13 Sep 2021 21:41
Operator : VA/SY
Sample : VSTDCCC050
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Sep 14 00:57:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.708	0.717	-1.3	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6