

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
 Data File : VD074264.D
 Acq On : 13 Sep 2022 14:20
 Operator : VA/SY
 Sample : N4453-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-T-5-(0-5)

Quant Time: Sep 14 01:55:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	98404	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	170491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	136277	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	57870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	59493	59.301	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	118.600%
35) Dibromofluoromethane	7.804	113	61446	54.128	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.260%
50) Toluene-d8	10.263	98	174633	41.786	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	83.580%
62) 4-Bromofluorobenzene	12.563	95	59821	39.536	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	79.080%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.787	84	17447	8.835	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
Data File : VD074264.D
Acq On : 13 Sep 2022 14:20
Operator : VA/SY
Sample : N4453-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-T-5-(0-5)

Quant Time: Sep 14 01:55:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

