

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
 Data File : VD074267.D
 Acq On : 13 Sep 2022 15:47
 Operator : VA/SY
 Sample : N4562-03
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-26524

Quant Time: Sep 14 01:55:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	2007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	5488	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.575	117	5353	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.498	152	2161	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	4862	237.615	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 475.240%#	
35) Dibromofluoromethane	7.799	113	3206	89.928	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 179.860%#	
50) Toluene-d8	10.269	98	6360	47.277	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 94.560%	
62) 4-Bromofluorobenzene	12.563	95	2875	62.990	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 125.980%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.781	84	2407	99.606	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
Data File : VD074267.D
Acq On : 13 Sep 2022 15:47
Operator : VA/SY
Sample : N4562-03
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-26524

Quant Time: Sep 14 01:55:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

