

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
 Data File : VD074271.D
 Acq On : 13 Sep 2022 17:44
 Operator : VA/SY
 Sample : N4582-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CRUSHED-MASONRY-PILE-7

Quant Time: Sep 14 01:56:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	106600	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	192498	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	175249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	82051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	60202	55.394	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.780%	
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	= 0.000%#	
50) Toluene-d8	10.263	98	246616	52.263	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.520%	
62) 4-Bromofluorobenzene	12.563	95	77050	46.232	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.460%	
Target Compounds						Qvalue
16) Acetone	4.022	43	3273	14.590	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
Data File : VD074271.D
Acq On : 13 Sep 2022 17:44
Operator : VA/SY
Sample : N4582-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CRUSHED-MASONRY-PILE-7

Quant Time: Sep 14 01:56:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

