

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
 Data File : VD074272.D
 Acq On : 13 Sep 2022 18:13
 Operator : VA/SY
 Sample : N4562-01RE
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHRT-2046

Quant Time: Sep 14 01:57:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	85771	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	148353	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	108383	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	38721	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.222	65	49630	56.756	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	113.520%	
35) Dibromofluoromethane	7.805	113	52915	53.532	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	107.060%	
50) Toluene-d8	10.263	98	169692	46.662	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	93.320%	
62) 4-Bromofluorobenzene	12.563	95	42743	31.027	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	62.060%	
Target Compounds						
20) Methylene Chloride	4.793	84	35541	29.890	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
Data File : VD074272.D
Acq On : 13 Sep 2022 18:13
Operator : VA/SY
Sample : N4562-01RE
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHRT-2046

Quant Time: Sep 14 01:57:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

