

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
 Data File : VD074277.D
 Acq On : 13 Sep 2022 20:37
 Operator : VA/SY
 Sample : N4591-01
 Misc : 5.08G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-091222

Quant Time: Sep 14 01:59:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	51074	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	94298	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	86692	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	40634	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	36711	70.502	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	141.000%	
35) Dibromofluoromethane	7.804	113	34676	55.299	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	110.600%	
50) Toluene-d8	10.263	98	111331	48.163	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	96.320%	
62) 4-Bromofluorobenzene	12.563	95	37077	45.273	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	90.540%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.798	84	16084	21.057	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091322\
Data File : VD074277.D
Acq On : 13 Sep 2022 20:37
Operator : VA/SY
Sample : N4591-01
Misc : 5.08G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-091222

Quant Time: Sep 14 01:59:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

