

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091323\
 Data File : VD077102.D
 Acq On : 13 Sep 2023 12:59
 Operator : JC/SY
 Sample : VD0913SBL02
 Misc : 5.00G/10.0ml/MSVOA_D/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VBLK148

Manual Integrations
 APPROVED

Reviewed By :Mahesh Dadoda 09/16/2023
 Supervised By :Semsettin Yesilyurt 09/18/2023

Quant Time: Sep 14 10:14:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM090123SMA.M
 Quant Title : SFAM01.0
 QLast Update : Sat Sep 02 01:35:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	287270	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	280769	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	130687	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.270	65	204835	28.937	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	115.760%
7) Chloroethane-d5	2.799	69	162146	27.856	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	111.440%
11) 1,1-Dichloroethene-d2	3.917	65	42494m	26.428	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	105.720%
21) 2-Butanone-d5	6.981	46	37854	42.450	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	84.900%
24) Chloroform-d	7.563	84	201788	25.671	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	102.680%
26) 1,2-Dichloroethane-d4	8.228	65	95279	25.574	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	102.280%
32) Benzene-d6	8.199	84	401766	25.908	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	103.640%
36) 1,2-Dichloropropane-d6	9.210	67	117540	26.430	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	105.720%
41) Toluene-d8	10.269	98	363615	25.615	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	102.480%
43) trans-1,3-Dichloroprop...	10.528	79	43049	21.245	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	85.000%
47) 2-Hexanone-d5	10.881	63	28698	40.404	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.800%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	92652	23.370	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.480%
66) 1,2-Dichlorobenzene-d4	13.816	152	113332	27.018	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	108.080%
Target Compounds						Qvalue
13) Acetone	4.028	43	6710	5.245	ug/L	100
16) Methylene chloride	4.787	84	31930m	5.461	ug/L	
25) Chloroform	7.593	83	5126	0.648	ug/L	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091323\
Data File : VD077102.D
Acq On : 13 Sep 2023 12:59
Operator : JC/SY
Sample : VD0913SBL02
Misc : 5.00G/10.0ml/MSVOA_D/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VBLK148

Manual Integrations
APPROVED

Reviewed By : Mahesh Dadoda 09/16/2023
Supervised By : Semsettin Yesilyurt 09/18/2023

Quant Time: Sep 14 10:14:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM090123SMA.M
Quant Title : SFAM01.0
QLast Update : Sat Sep 02 01:35:39 2023
Response via : Initial Calibration

