

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091324\
 Data File : VD079837.D
 Acq On : 13 Sep 2024 16:25
 Operator : RP/MD
 Sample : P3993-08
 Misc : 4.96G/10ml/MSVOA_D/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0UF6

Quant Time: Sep 13 23:52:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
 Quant Title : SFAM01.0
 QLast Update : Fri Sep 13 23:49:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	205137	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	201571	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	76656	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	105886	11.870	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	47.480%
7) Chloroethane-d5	2.799	69	103139	14.604	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	58.400%
11) 1,1-Dichloroethene-d2	3.922	65	20309	12.540	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	50.160%
21) 2-Butanone-d5	6.987	46	32953	39.761	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	79.520%
24) Chloroform-d	7.569	84	116578	17.288	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	69.160%
26) 1,2-Dichloroethane-d4	8.228	65	67385	19.110	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	76.440%
32) Benzene-d6	8.199	84	224169	15.752	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	63.000%
36) 1,2-Dichloropropane-d6	9.210	67	80017	18.359	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	73.440%
41) Toluene-d8	10.269	98	193390	15.443	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	61.760%
43) trans-1,3-Dichloroprop...	10.522	79	31934	17.042	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	68.160%
47) 2-Hexanone-d5	10.875	63	27736	39.038	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.080%
56) 1,1,2,2-Tetrachloroeth...	12.645	84	62657	19.563	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	78.240%
66) 1,2-Dichlorobenzene-d4	13.810	152	62315	21.556	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	86.240%
Target Compounds						
13) Acetone	4.022	43	13760	13.618	ug/L	Qvalue 61

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091324\
Data File : VD079837.D
Acq On : 13 Sep 2024 16:25
Operator : RP/MD
Sample : P3993-08
Misc : 4.96G/10ml/MSVOA_D/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0UF6

Quant Time: Sep 13 23:52:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM091024SMA.M
Quant Title : SFAM01.0
QLast Update : Fri Sep 13 23:49:48 2024
Response via : Initial Calibration

