

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070347.D
 Acq On : 14 Sep 2021 12:17
 Operator : VA/SY
 Sample : M3719-03RE
 Misc : 5.07G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-091021RE

Quant Time: Sep 15 02:45:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	118945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	231141	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	310500	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	151294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	111303	72.877	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	145.760%	
35) Dibromofluoromethane	7.914	113	102937	62.565	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	125.120%	
50) Toluene-d8	10.337	98	315040	52.738	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	105.480%	
62) 4-Bromofluorobenzene	12.626	95	140214	67.162	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	134.320%	
Target Compounds						
16) Acetone	4.155	43	3084	11.310	ug/l	Qvalue # 36

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
Data File : VD070347.D
Acq On : 14 Sep 2021 12:17
Operator : VA/SY
Sample : M3719-03RE
Misc : 5.07G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-091021RE

Quant Time: Sep 15 02:45:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

