

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070349.D
 Acq On : 14 Sep 2021 13:14
 Operator : VA/SY
 Sample : M3718-01RE
 Misc : 5.09G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-03-091021RE

Quant Time: Sep 15 02:45:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	28299	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	47080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	64892	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	31024	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	25462	70.073	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 140.140%	
35) Dibromofluoromethane	7.909	113	20323	60.644	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 121.280%	
50) Toluene-d8	10.338	98	64425	52.949	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.900%	
62) 4-Bromofluorobenzene	12.626	95	29072	68.367	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 136.740%	
Target Compounds						
16) Acetone	4.162	43	1968	30.336	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
Data File : VD070349.D
Acq On : 14 Sep 2021 13:14
Operator : VA/SY
Sample : M3718-01RE
Misc : 5.09G/5.00ml/MSVOA_D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-03-091021RE

Quant Time: Sep 15 02:45:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

