

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070352.D
 Acq On : 14 Sep 2021 14:40
 Operator : VA/SY
 Sample : M3687-23RE
 Misc : 7.21G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A010015RE

Quant Time: Sep 15 02:46:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	297059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	596816	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	758098	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	342883	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	204976	53.739	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.480%
35) Dibromofluoromethane	7.909	113	220761	51.966	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.940%
50) Toluene-d8	10.338	98	826011	53.553	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.100%
62) 4-Bromofluorobenzene	12.626	95	395311	73.334	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	146.660%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
Data File : VD070352.D
Acq On : 14 Sep 2021 14:40
Operator : VA/SY
Sample : M3687-23RE
Misc : 7.21G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A010015RE

Quant Time: Sep 15 02:46:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

