

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070354.D
 Acq On : 14 Sep 2021 15:38
 Operator : VA/SY
 Sample : M3724-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleID :
 31190

Quant Time: Sep 15 02:46:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	16453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	27380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	31726	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	14095	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	14376	68.049	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.100%
35) Dibromofluoromethane	7.914	113	9838	50.479	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.960%
50) Toluene-d8	10.332	98	33808	47.777	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	95.560%
62) 4-Bromofluorobenzene	12.626	95	13833	55.936	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	111.880%
Target Compounds						
52) Toluene	10.402	92	6376	12.304	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070354.D
 Acq On : 14 Sep 2021 15:38
 Operator : VA/SY
 Sample : M3724-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 31190

Quant Time: Sep 15 02:46:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

