

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
 Data File : VD070359.D
 Acq On : 14 Sep 2021 18:01
 Operator : VA/SY
 Sample : M3687-30
 Misc : 3.23G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 A003015

Quant Time: Sep 15 02:47:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
 Quant Title : SW846 8260
 QLast Update : Thu Sep 09 03:07:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	318865	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	632759	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	780687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	368185	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	198474	48.476	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.960%
35) Dibromofluoromethane	7.914	113	221598	49.200	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.400%
50) Toluene-d8	10.338	98	849862	51.969	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.940%
62) 4-Bromofluorobenzene	12.626	95	365572	63.965	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	127.920%
Target Compounds						
32) 1,1,1-Trichloroethane	7.903	97	8750	1.227	ug/l #	46
64) Tetrachloroethene	10.867	164	5404	1.137	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091421\
Data File : VD070359.D
Acq On : 14 Sep 2021 18:01
Operator : VA/SY
Sample : M3687-30
Misc : 3.23G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
A003015

Quant Time: Sep 15 02:47:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090821S.M
Quant Title : SW846 8260
QLast Update : Thu Sep 09 03:07:09 2021
Response via : Initial Calibration

