

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091422\
 Data File : VD074287.D
 Acq On : 14 Sep 2022 16:19
 Operator : VA/SY
 Sample : N4579-15
 Misc : 2.86G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 CHERRY-HILL-COMP-14

Quant Time: Sep 15 04:14:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	98671	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	172201	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	237156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	166626	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	24771	24.624	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	49.240%#
35) Dibromofluoromethane	7.805	113	40355	33.961	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	67.920%
50) Toluene-d8	10.263	98	88311	20.921	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	41.840%#
62) 4-Bromofluorobenzene	12.563	95	68329	45.762	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	91.520%
Target Compounds						Qvalue
20) Methylene Chloride	4.793	84	18823	10.031	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091422\
Data File : VD074287.D
Acq On : 14 Sep 2022 16:19
Operator : VA/SY
Sample : N4579-15
Misc : 2.86G/5.00ml/MSVOA_D/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
CHERRY-HILL-COMP-14

Quant Time: Sep 15 04:14:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

