

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091422\
 Data File : VD074294.D
 Acq On : 14 Sep 2022 19:40
 Operator : VA/SY
 Sample : N4538-05
 Misc : 5.74G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 MBD-MP-1-1-2

Quant Time: Sep 15 04:15:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	105838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	189780	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	158614	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	68376	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	39340	36.459	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	72.920%
35) Dibromofluoromethane	7.798	113	42393	32.206	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	64.420%
50) Toluene-d8	10.269	98	151881	32.648	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	65.300%
62) 4-Bromofluorobenzene	12.563	95	44757	23.939	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	47.880%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.799	84	17859	8.074	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091422\
Data File : VD074294.D
Acq On : 14 Sep 2022 19:40
Operator : VA/SY
Sample : N4538-05
Misc : 5.74G/5.00ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
MBD-MP-1-1-2

Quant Time: Sep 15 04:15:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

