

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074305.D
 Acq On : 15 Sep 2022 13:43
 Operator : VA/SY
 Sample : N4591-01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OR-02-091222

Quant Time: Sep 15 14:22:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	116590	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	205942	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	180158	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	76254	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	53310	44.849	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	89.700%
35) Dibromofluoromethane	7.805	113	52509	37.260	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.520%
50) Toluene-d8	10.269	98	209251	41.450	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	82.900%
62) 4-Bromofluorobenzene	12.569	95	59034	30.829	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	61.660%
Target Compounds						
16) Acetone	4.016	43	1793	7.308	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074305.D
Acq On : 15 Sep 2022 13:43
Operator : VA/SY
Sample : N4591-01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OR-02-091222

Quant Time: Sep 15 14:22:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

