

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074310.D
 Acq On : 15 Sep 2022 16:05
 Operator : VA/SY
 Sample : N4573-02RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 72-11998RE

Quant Time: Sep 16 05:54:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	2516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	4540	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	4470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.504	152	2198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	2136	83.272	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 166.540%#	
35) Dibromofluoromethane	7.804	113	1714	56.868	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 113.740%	
50) Toluene-d8	10.269	98	5513	49.538	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 99.080%	
62) 4-Bromofluorobenzene	12.563	95	1864	47.630	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 95.260%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074310.D
Acq On : 15 Sep 2022 16:05
Operator : VA/SY
Sample : N4573-02RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
72-11998RE

Quant Time: Sep 16 05:54:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

