

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074317.D
 Acq On : 15 Sep 2022 19:22
 Operator : VA/SY
 Sample : N4610-01
 Misc : 6.28G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EP-BOTTOM

Quant Time: Sep 16 05:55:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.869	168	18165	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	34572	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	32598	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.510	152	13245	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	17498	94.484	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 188.960%#	
35) Dibromofluoromethane	7.810	113	13189	57.501	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 115.000%	
50) Toluene-d8	10.263	98	38050	44.899	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 89.800%	
62) 4-Bromofluorobenzene	12.569	95	12330	40.319	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	= 80.640%	
Target Compounds						Qvalue
16) Acetone	4.010	43	1379	36.075	ug/l #	42
20) Methylene Chloride	4.793	84	4053	12.904	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074317.D
Acq On : 15 Sep 2022 19:22
Operator : VA/SY
Sample : N4610-01
Misc : 6.28G/5.00ml/MSVOA_D/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EP-BOTTOM

Quant Time: Sep 16 05:55:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

