

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074318.D
 Acq On : 15 Sep 2022 19:50
 Operator : VA/SY
 Sample : N4584-01
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 33811

Quant Time: Sep 16 05:55:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	27065	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	53117	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	53039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	25243	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	25058	90.812	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	181.620%#
35) Dibromofluoromethane	7.805	113	20094	56.990	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.980%
50) Toluene-d8	10.269	98	69390	53.292	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.580%
62) 4-Bromofluorobenzene	12.569	95	24204	53.744	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	=	107.480%
Target Compounds						
						Qvalue
16) Acetone	4.028	43	1347	23.650	ug/l	100
20) Methylene Chloride	4.793	84	8861	22.166	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074318.D
Acq On : 15 Sep 2022 19:50
Operator : VA/SY
Sample : N4584-01
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
33811

Quant Time: Sep 16 05:55:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

