

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
 Data File : VD074319.D
 Acq On : 15 Sep 2022 20:19
 Operator : VA/SY
 Sample : N4627-01
 Misc : 5.87G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-3-091422

Quant Time: Sep 16 05:55:56 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
 Quant Title : SW846 8260
 QLast Update : Sat Sep 10 04:45:19 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	109268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	197975	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.575	117	174832	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.510	152	77052	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	60925	54.690	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.380%	
35) Dibromofluoromethane	7.804	113	62165	46.704	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 93.400%	
50) Toluene-d8	10.269	98	245557	50.599	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.200%	
62) 4-Bromofluorobenzene	12.569	95	69765	39.742	ug/l	0.01
Spiked Amount	50.000	Range	25 - 144	Recovery	= 79.480%	
Target Compounds						
20) Methylene Chloride	4.793	84	21014	10.168	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD091522\
Data File : VD074319.D
Acq On : 15 Sep 2022 20:19
Operator : VA/SY
Sample : N4627-01
Misc : 5.87G/5.00ml/MSVOA_D/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-3-091422

Quant Time: Sep 16 05:55:56 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D090922S.M
Quant Title : SW846 8260
QLast Update : Sat Sep 10 04:45:19 2022
Response via : Initial Calibration

